

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: CO-8R-WCDUP	SDG No.: Q1929 Sample ID: Q1907-01 Percent Solids for Spike Sample: 84.6
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/Kg	+/-20	2480		2500		1	0.52		05/06/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q1929
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1907-01
Client ID:	CO-8R-WCMSD	Percent Solids for Spike Sample:	84.6

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/Kg	+/-20	2480		2620		1			05/06/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-12-S-202504DUP	SDG No.: Q1929 Sample ID: Q1913-01 Percent Solids for Spike Sample: 80.5
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		05/05/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q1929
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1916-04
Client ID:	WC-12DUP	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Reactive Cyanide	mg/Kg	+/-20	0.0083	U	0.0087	J	1	200	*	05/02/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: MH-QDUP	SDG No.: Q1929 Sample ID: Q1928-01 Percent Solids for Spike Sample: 87.1
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Paint Filter	ml/100gm	+/-20	1.00	U	1.00	U	1	0		05/02/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: MH-QDUP	SDG No.: Q1929 Sample ID: Q1928-04 Percent Solids for Spike Sample: 100
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Reactive Sulfide	mg/Kg	+/-20	1.60	J	1.60	J	1	0		05/02/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-AA-02-CDUP	SDG No.: Q1929 Sample ID: Q1929-02 Percent Solids for Spike Sample: 86.7
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	12.4		12.4		1	0.08		05/01/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-A4-02-CDUP	SDG No.: Q1929 Sample ID: Q1929-03 Percent Solids for Spike Sample: 100
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Corrosivity	pH	+/-20	12.4		12.4		1	0.08		05/01/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-A4-02-CDUP	SDG No.: Q1929 Sample ID: Q1929-04 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
ASTM Ammonia	mg/L	+/-20	0.28		0.28		1	0		05/06/2025
ASTM COD	mg/L	+/-20	20.5		21.5		1	4.76		05/06/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-A4-02-CMSD	SDG No.: Q1929 Sample ID: Q1929-04 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
ASTM Ammonia	mg/L	+/-20	1.00		1.10		1	10		05/06/2025
ASTM COD	mg/L	+/-20	65.3		67.3		1	3.02		05/06/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-AA-04-CDUP	SDG No.: Q1929 Sample ID: Q1929-10 Percent Solids for Spike Sample: 81.5
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TS	%	+/-5	81.9		83.3		1	1.69		05/01/2025
TVS	%	+/-5	4.10	J	4.00	J	1	2.47		05/01/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q1929
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1929-12
Client ID:	WC-A1-04-CDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
ASTM TS	mg/L	+/-5	1110		1130		1	1.43		05/05/2025
ASTM Oil and Grease	mg/L	+/-18	0.40	J	0.40	J	1	0		05/05/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q1929
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1929-12
Client ID:	WC-A1-04-CMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
ASTM Oil and Grease	mg/L	+/-18	20.6		20.5		1	0.49		05/05/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-A1-07-CDUP	SDG No.: Q1929 Sample ID: Q2001-14 Percent Solids for Spike Sample: 100
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	12.3		12.3		1	0.08		05/12/2025