

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1929

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135661

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135661BS							
ASTM Oil and Grease	mg/L	20.0	16.8		84	1	78-114	05/05/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1929

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135676

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135676BS							
Oil and Grease	mg/Kg	100	95.0		95	1	80-120	05/06/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1929

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135682

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135682BS							
ASTM COD	mg/L	50	50.0		100	1	90-110	05/06/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1929

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135677

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167867BS							
ASTM Ammonia	mg/L	1	0.98		98	1	90-110	05/06/2025