



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q1929
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1929 **SAS No.:** Q1929
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	97	40 - 160	CV	05/05/2025	11:33	LB135662
CRI01	Arsenic	20.6	20.0	103	40 - 160	P	05/05/2025	14:47	LB135674
	Barium	92.2	100	92	40 - 160	P	05/05/2025	14:47	LB135674
	Cadmium	5.79	6.0	96	40 - 160	P	05/05/2025	14:47	LB135674
	Chromium	10.3	10.0	103	40 - 160	P	05/05/2025	14:47	LB135674
	Copper	22.4	20.0	112	40 - 160	P	05/05/2025	14:47	LB135674
	Lead	11.2	12.0	93	40 - 160	P	05/05/2025	14:47	LB135674
	Nickel	39.9	40.0	100	40 - 160	P	05/05/2025	14:47	LB135674
	Selenium	22.8	20.0	114	40 - 160	P	05/05/2025	14:47	LB135674
	Silver	10.3	10.0	103	40 - 160	P	05/05/2025	14:47	LB135674
	Zinc	42.7	40.0	107	40 - 160	P	05/05/2025	14:47	LB135674