



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q1929				
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q1929	SAS No.:	Q1929		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB114	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	11:18	LB135662	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q1929

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1929

SAS No.: Q1929

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB71	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	11:30	LB135662
CCB72	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	12:03	LB135662
CCB73	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	12:38	LB135662
CCB74	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	13:06	LB135662
CCB75	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	13:43	LB135662
CCB76	Mercury	0.20	+/-0.20	U	0.20	CV	05/05/2025	13:55	LB135662

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1929</u>							
Contract: <u>ENTA05</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1929</u>		SAS No.: <u>Q1929</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	14:41	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	14:41	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	14:41	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	14:41	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	14:41	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	14:41	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	14:41	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	14:41	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	14:41	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	14:41	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q1929							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q1929	SAS No.: Q1929						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	15:26	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	15:26	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	15:26	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	15:26	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	15:26	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	15:26	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	15:26	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	15:26	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	15:26	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	15:26	LB135674
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	16:12	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	16:12	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	16:12	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	16:12	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	16:12	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	16:12	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	16:12	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	16:12	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	16:12	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	16:12	LB135674
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	17:00	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	17:00	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	17:00	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	17:00	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	17:00	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	17:00	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	17:00	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	17:00	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	17:00	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	17:00	LB135674
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	17:59	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	17:59	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	17:59	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	17:59	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	17:59	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	17:59	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	17:59	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	17:59	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	17:59	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT			SDG No.:	Q1929				
Contract:	ENTA05	Lab Code:	CHEM	Case No.:	Q1929	SAS No.:	Q1929		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	17:59	LB135674
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	18:47	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	18:47	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	18:47	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	18:47	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	18:47	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	18:47	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	18:47	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	18:47	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	18:47	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	18:47	LB135674
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	19:37	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	19:37	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	19:37	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	19:37	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	19:37	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	19:37	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	19:37	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	19:37	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	19:37	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	19:37	LB135674
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	20:23	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	20:23	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	20:23	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	20:23	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	20:23	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	20:23	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	20:23	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	20:23	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	20:23	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	20:23	LB135674
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	21:12	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	21:12	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	21:12	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	21:12	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	21:12	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	21:12	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	21:12	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	21:12	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1929</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1929</u>	SAS No.: <u>Q1929</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	21:12	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	21:12	LB135674
CCB09	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	22:03	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	22:03	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	22:03	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	22:03	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	22:03	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	22:03	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	22:03	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	22:03	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	22:03	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	22:03	LB135674
CCB10	Arsenic	20.0	+/-20.0	U	20.0	P	05/05/2025	22:37	LB135674
	Barium	100	+/-100	U	100	P	05/05/2025	22:37	LB135674
	Cadmium	6.00	+/-6.00	U	6.00	P	05/05/2025	22:37	LB135674
	Chromium	10.0	+/-10.0	U	10.0	P	05/05/2025	22:37	LB135674
	Copper	20.0	+/-20.0	U	20.0	P	05/05/2025	22:37	LB135674
	Lead	12.0	+/-12.0	U	12.0	P	05/05/2025	22:37	LB135674
	Nickel	40.0	+/-40.0	U	40.0	P	05/05/2025	22:37	LB135674
	Selenium	20.0	+/-20.0	U	20.0	P	05/05/2025	22:37	LB135674
	Silver	10.0	+/-10.0	U	10.0	P	05/05/2025	22:37	LB135674
	Zinc	40.0	+/-40.0	U	40.0	P	05/05/2025	22:37	LB135674

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT **SDG No.:** Q1929
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1929 **SAS No.:** Q1929

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------