

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** Q1929
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1929 **SAS No.:** Q1929
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.47			-20	20	05/05/2025	14:52	LB135674
	Barium	0.40	6.0	7	-94	106	05/05/2025	14:52	LB135674
	Cadmium	-1.99	1.0	199	-5	7	05/05/2025	14:52	LB135674
	Chromium	53.8	52.0	104	42	62	05/05/2025	14:52	LB135674
	Copper	12.1	2.0	605	-18	22	05/05/2025	14:52	LB135674
	Lead	-1.42			-12	12	05/05/2025	14:52	LB135674
	Nickel	1.16	2.0	58	-38	42	05/05/2025	14:52	LB135674
	Selenium	-15.2			-20	20	05/05/2025	14:52	LB135674
	Silver	0.79			-10	10	05/05/2025	14:52	LB135674
	Zinc	3.23			-40	40	05/05/2025	14:52	LB135674
ICSAB01	Arsenic	112	104	108	88.4	120	05/05/2025	15:07	LB135674
	Barium	496	537	92	437	637	05/05/2025	15:07	LB135674
	Cadmium	1050	972	108	826	1120	05/05/2025	15:07	LB135674
	Chromium	572	542	106	460	624	05/05/2025	15:07	LB135674
	Copper	520	511	102	434	588	05/05/2025	15:07	LB135674
	Lead	48.4	49.0	99	37	61	05/05/2025	15:07	LB135674
	Nickel	1020	954	107	810	1100	05/05/2025	15:07	LB135674
	Selenium	35.7	46.0	78	26	66	05/05/2025	15:07	LB135674
	Silver	195	201	97	170	232	05/05/2025	15:07	LB135674
	Zinc	1070	952	112	809	1095	05/05/2025	15:07	LB135674