

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1929
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1929 **SAS No.:** Q1929

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167837BS							
Arsenic	ug/L	4000	3730		93	80 - 120	P
Barium	ug/L	1000	850		85	80 - 120	P
Cadmium	ug/L	1000	922		92	80 - 120	P
Chromium	ug/L	2000	1830		92	80 - 120	P
Copper	ug/L	1500	1410		94	80 - 120	P
Lead	ug/L	5000	4570		91	80 - 120	P
Nickel	ug/L	2500	2280		91	80 - 120	P
Selenium	ug/L	10000	9480		95	80 - 120	P
Silver	ug/L	380	339		89	80 - 120	P
Zinc	ug/L	1000	916		92	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1929
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1929 **SAS No.:** Q1929

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167839BS Mercury	ug/L	4.0	4.05		101	80 - 120	CV