

Surrogate Summary

SDG No.: Q1929

Client: ENTACT

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP071388.D	PIBLK-PP071388.D	Tetrachloro-m-xylene	1	20	17.1	85		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.6	88		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	17.2	86		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.6	88		70 (60)	130 (140)
I.BLK-PP071735.D	PIBLK-PP071735.D	Tetrachloro-m-xylene	1	20	16.7	83		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.2	91		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	17.8	89		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.1	101		70 (60)	130 (140)
PB167834BL	PB167834BL	Tetrachloro-m-xylene	1	20	20.1	100		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.6	108		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.8	104		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	23.8	119		30 (32)	150 (175)
PB167834BS	PB167834BS	Tetrachloro-m-xylene	1	20	20.7	103		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	22.1	111		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.0	100		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	23.9	119		30 (32)	150 (175)
Q1928-01MS	MH-QMS	Tetrachloro-m-xylene	1	20	20.5	102		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.3	97		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.9	110		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.0	110		30 (32)	150 (175)
Q1928-01MSD	MH-QMSD	Tetrachloro-m-xylene	1	20	20.2	101		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.5	97		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.1	101		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.0	105		30 (32)	150 (175)
Q1929-02	WC-A4-02-C	Tetrachloro-m-xylene	1	20	14.6	73		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	13.5	68		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.0	85		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	16.3	82		30 (32)	150 (175)
Q1929-06	WC-A1-03-C	Tetrachloro-m-xylene	1	20	13.7	69		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	11.6	58		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.9	90		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	13.6	68		30 (32)	150 (175)
Q1929-10	WC-A1-04-C	Tetrachloro-m-xylene	1	20	22.3	111		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	15.1	76		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.8	89		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	17.8	89		30 (32)	150 (175)
I.BLK-PP071750.D	PIBLK-PP071750.D	Tetrachloro-m-xylene	1	20	16.9	85		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.6	93		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	18.8	94		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.3	101		70 (60)	130 (140)