



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1929

Client: ENTACT

Analytical Method: **8082A** **Datafile :** PP071738.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167834BS	AR1016	166.5	157	ug/kg	94				40 (71)	140 (120)	
	AR1260	166.5	165	ug/kg	99				40 (65)	140 (130)	