

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

B-167-SB01MS

Contract: ENTA05

Lab Code: CHEM Case No.: Q1929

SAS No.: Q1929 SDG NO.: Q1929

Lab Sample ID: Q1901-08MS

Date(s) Analyzed: 05/06/2025 05/06/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.01	7.96	8.06	64.2	0.8
	2	8.65	8.60	8.70	63.7	
2,4,5-TP(Silvex)	1	8.86	8.81	8.91	60.9	4
	2	9.53	9.48	9.58	63.4	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

B-167-SB01MSD

Contract: ENTA05

Lab Code: CHEM Case No.: Q1929

SAS No.: Q1929

SDG NO.: Q1929

Lab Sample ID: Q1901-08MSD

Date(s) Analyzed: 05/06/2025 05/06/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.01	7.96	8.06	67.0	1.1
	2	8.65	8.60	8.70	66.3	
2,4,5-TP(Silvex)	1	8.86	8.81	8.91	63.4	4.2
	2	9.53	9.48	9.58	66.1	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB167871BS

Contract: ENTA05

Lab Code: CHEM Case No.: Q1929

SAS No.: Q1929 SDG NO.: Q1929

Lab Sample ID: PB167871BS

Date(s) Analyzed: 05/06/2025 05/06/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4,5-TP(Silvex)	1	8.86	8.81	8.91	5.30	1.9
	2	9.53	9.48	9.58	5.40	
2,4-D	1	8.01	7.96	8.06	5.20	1.9
	2	8.65	8.60	8.70	5.30	