

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q1929 **SAS No.:** Q1929 **SDG NO.:** Q1929

Continuing Calib Date: 05/06/2025 **Initial Calibration Date(s):** 04/23/2025 04/23/2025

Continuing Calib Time: 09:50 **Initial Calibration Time(s):** 11:01 12:37

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	6.94	6.94	6.84	7.04	0.00
2,4-D	8.02	8.02	7.92	8.12	0.00
2,4,5-TP(Silvex)	8.86	8.86	8.76	8.96	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q1929 **SAS No.:** Q1929 **SDG NO.:** Q1929

Continuing Calib Date: 05/06/2025 **Initial Calibration Date(s):** 04/23/2025 04/23/2025

Continuing Calib Time: 09:50 **Initial Calibration Time(s):** 11:01 12:37

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.46	7.47	7.37	7.57	0.01
2,4-D	8.65	8.66	8.56	8.76	0.01
2,4,5-TP(Silvex)	9.53	9.54	9.44	9.64	0.01