

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** ENTA05

**Lab Code:** CHEM      **Case No.:** Q1929      **SAS No.:** Q1929      **SDG NO.:** Q1929

**Continuing Calib Date:** 05/06/2025      **Initial Calibration Date(s):** 04/23/2025      04/23/2025

**Continuing Calib Time:** 14:15      **Initial Calibration Time(s):** 11:01      12:37

**GC Column:** RTX-CLP      **ID:** 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 6.94       | 6.94      | 6.84      | 7.04 | 0.00       |
| 2,4-D            | 8.02       | 8.02      | 7.92      | 8.12 | 0.00       |
| 2,4,5-TP(Silvex) | 8.86       | 8.86      | 8.76      | 8.96 | 0.00       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** ENTA05

**Lab Code:** CHEM      **Case No.:** Q1929      **SAS No.:** Q1929      **SDG NO.:** Q1929

**Continuing Calib Date:** 05/06/2025      **Initial Calibration Date(s):** 04/23/2025      04/23/2025

**Continuing Calib Time:** 14:15      **Initial Calibration Time(s):** 11:01      12:37

**GC Column:** RTX-CLP2      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.46       | 7.47      | 7.37      | 7.57 | 0.01       |
| 2,4-D            | 8.65       | 8.66      | 8.56      | 8.76 | 0.01       |
| 2,4,5-TP(Silvex) | 9.53       | 9.54      | 9.44      | 9.64 | 0.01       |