

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG NO.: Q1929

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL01 Date Analyzed: 05/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030052.D Time Analyzed: 09:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.857	8.762	8.962	695.180	712.500	-2.4
2,4-D	8.015	7.919	8.119	682.930	705.000	-3.1
2,4-DCAA	6.938	6.842	7.042	709.030	750.000	-5.5

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG NO.: Q1929

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL01 Date Analyzed: 05/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030052.D Time Analyzed: 09:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.531	9.442	9.642	717.130	712.500	0.6
2,4-D	8.651	8.561	8.761	682.800	705.000	-3.1
2,4-DCAA	7.459	7.367	7.567	734.700	750.000	-2.0