

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG NO.: Q1929

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 05/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030059.D Time Analyzed: 14:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	8.856	8.762	8.962	744.780	712.500	4.5
2,4-D	8.015	7.919	8.119	719.560	705.000	2.1
2,4-DCAA	6.938	6.842	7.042	763.780	750.000	1.8

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG NO.: Q1929

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 05/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030059.D Time Analyzed: 14:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.532	9.442	9.642	772.440	712.500	8.4
2,4-D	8.652	8.561	8.761	732.170	705.000	3.9
2,4-DCAA	7.461	7.367	7.567	793.080	750.000	5.7