

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG NO.: Q1929

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/05/2025

Lab Sample No.: PSTDCCC050 Data File : PD088408.D Time Analyzed: 08:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.073	8.975	9.175	50.200	50.000	0.4
Endrin	6.575	6.476	6.676	54.180	50.000	8.4
gamma-BHC (Lindane)	4.331	4.232	4.432	56.770	50.000	13.5
Heptachlor	4.930	4.831	5.031	55.420	50.000	10.8
Heptachlor epoxide	5.691	5.592	5.792	55.740	50.000	11.5
Methoxychlor	7.493	7.395	7.595	49.590	50.000	-0.8
Tetrachloro-m-xylene	3.550	3.452	3.652	54.530	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG NO.: Q1929

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 05/05/2025

Lab Sample No.: PSTDCCC050 Data File : PD088408.D Time Analyzed: 08:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.074	7.977	8.177	44.490	50.000	-11.0
Endrin	5.791	5.693	5.893	47.170	50.000	-5.7
gamma-BHC (Lindane)	3.731	3.633	3.833	48.880	50.000	-2.2
Heptachlor	4.084	3.986	4.186	46.430	50.000	-7.1
Heptachlor epoxide	4.874	4.777	4.977	49.060	50.000	-1.9
Methoxychlor	6.756	6.659	6.859	41.030	50.000	-17.9
Tetrachloro-m-xylene	2.882	2.783	2.983	49.100	50.000	-1.8