

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1929

Client: ENTACT

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1916-04MSD	Client Sample ID:	WC-12MSD					DataFile:	BF142285.D		
Pyridine	500	0	0	ug/L	0	*	0		20 (10)	160 (109)	20 (20)
1,4-Dichlorobenzene	500	0	390	ug/L	78		3		70 (55)	130 (125)	20 (20)
2-Methylphenol	500	0	430	ug/L	86		2		70 (60)	130 (131)	20 (20)
3+4-Methylphenols	500	0	400	ug/L	80		2		20 (54)	160 (136)	20 (20)
Hexachloroethane	500	0	400	ug/L	80		2		20 (19)	160 (146)	20 (20)
Nitrobenzene	500	0	470	ug/L	94		0		70 (62)	130 (112)	20 (20)
Hexachlorobutadiene	500	0	410	ug/L	82		0		70 (52)	130 (125)	20 (20)
2,4,6-Trichlorophenol	500	0	480	ug/L	96		0		70 (78)	130 (112)	20 (20)
2,4,5-Trichlorophenol	500	0	490	ug/L	98		2		70 (71)	130 (111)	20 (20)
2,4-Dinitrotoluene	500	0	550	ug/L	110		2		70 (74)	130 (137)	20 (20)
Hexachlorobenzene	500	0	450	ug/L	90		2		70 (72)	130 (115)	20 (20)
Pentachlorophenol	1000	0	1000	ug/L	100		0		20 (52)	160 (162)	20 (20)