

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG NO.: Q1929

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/05/2025

Lab File ID: BM050096.D Time Analyzed: 12:18

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	253723	7.745	967980	10.54	663376	14.40
UPPER LIMIT	507446	8.245	1935960	11.039	1326750	14.898
LOWER LIMIT	126862	7.245	483990	10.039	331688	13.898
EPA SAMPLE NO.						
01 PB167815TB	249102	7.75	868897	10.55	612335	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.