

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG NO.: Q1929

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/05/2025

Lab File ID: BP024522.D Time Analyzed: 12:38

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	132437	7.705	539349	10.48	358648	14.34
UPPER LIMIT	264874	8.205	1078700	10.975	717296	14.84
LOWER LIMIT	66218.5	7.205	269675	9.975	179324	13.84
EPA SAMPLE NO.						
01 WC-A1-03-C	134404	7.71	525227	10.48	306807	14.34
02 WC-A1-04-C	145173	7.71	561713	10.48	359614	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.