

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG NO.: Q1929

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/06/2025

Lab File ID: BP024540.D Time Analyzed: 10:17

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	159873	7.705	632107	10.48	387042	14.34
UPPER LIMIT	319746	8.205	1264210	10.975	774084	14.839
LOWER LIMIT	79936.5	7.205	316054	9.975	193521	13.839
EPA SAMPLE NO.						
01 PB167847BL	167503	7.71	646551	10.48	376840	14.34
02 PB167847BS	173185	7.70	691867	10.48	422844	14.34

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.