

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q1929 SAS No.: Q1929 SDG No.: Q1929

Instrument ID: MSVOA_N Calibration Date/Time: 05/05/2025 11:32

Lab File ID: VN086478.D Init. Calib. Date(s): 04/15/2025 04/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:29 14:29

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.822	0.759		-7.66	20
1,1-Dichloroethene	0.588	0.557		-5.27	20
2-Butanone	0.451	0.469		3.99	20
Carbon Tetrachloride	0.441	0.442		0.23	20
Chloroform	1.180	1.306		10.68	20
Benzene	1.485	1.541		3.77	20
1,2-Dichloroethane	0.474	0.533		12.45	20
Trichloroethene	0.352	0.372		5.68	20
Tetrachloroethene	0.385	0.386		0.26	20
Chlorobenzene	1.116	1.173	0.3	5.11	20
1,2-Dichloroethane-d4	0.725	0.817		12.69	20
Dibromofluoromethane	0.232	0.264		13.79	20
Toluene-d8	1.240	1.280		3.23	20
4-Bromofluorobenzene	0.452	0.478		5.75	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.