

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1929

Client: ENTACT

Analytical Method: SW8260-Low

Datafile : VN086482.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0505WBSD01	Vinyl chloride	20	18.8	ug/L	94	1		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	18.9	ug/L	95	1		70 (74)	130 (110)	20 (20)
	2-Butanone	100	96.5	ug/L	97	10		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	21.0	ug/L	105	2		70 (77)	130 (113)	20 (20)
	Chloroform	20	20.8	ug/L	104	3		70 (79)	130 (113)	20 (20)
	Benzene	20	19.9	ug/L	100	2		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	21.7	ug/L	109	6		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	21.0	ug/L	105	3		70 (77)	130 (113)	20 (20)
	Tetrachloroethene	20	21.1	ug/L	106	6		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	20.9	ug/L	104	1		70 (82)	130 (109)	20 (20)