



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Verification

**Client:** VERINA CONSULTING GROUP, LLC

**SDG No.:** Q1930

**Project:** Rotor Clip NJ WTD - 2025

**RunNo.:** LB135627

| Analyte                                     | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|---------------------------------------------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID: <b>ICV</b><br>Residual Chlorine  | mg/L  | 0.416  | 0.4        | 104           | 90-110                    | 05/01/2025       |
| Sample ID: <b>CCV1</b><br>Residual Chlorine | mg/L  | 0.406  | 0.4        | 102           | 90-110                    | 05/01/2025       |
| Sample ID: <b>CCV2</b><br>Residual Chlorine | mg/L  | 0.395  | 0.4        | 99            | 90-110                    | 05/01/2025       |

## Initial and Continuing Calibration Verification

**Client:** VERINA CONSULTING GROUP, LLC

**SDG No.:** Q1930

**Project:** Rotor Clip NJ WTD - 2025

**RunNo.:** LB135665

| Analyte                                | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|----------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV1</b><br>Ammonia as N | mg/L  | 0.95   | 1          | 95         | 90-110                 | 05/05/2025    |
| Sample ID: <b>CCV1</b><br>Ammonia as N | mg/L  | 0.96   | 1          | 96         | 90-110                 | 05/05/2025    |
| Sample ID: <b>CCV2</b><br>Ammonia as N | mg/L  | 0.98   | 1          | 98         | 90-110                 | 05/05/2025    |
| Sample ID: <b>CCV3</b><br>Ammonia as N | mg/L  | 0.99   | 1          | 99         | 90-110                 | 05/05/2025    |
| Sample ID: <b>CCV4</b><br>Ammonia as N | mg/L  | 0.98   | 1          | 98         | 90-110                 | 05/05/2025    |

## Initial and Continuing Calibration Verification

**Client:** VERINA CONSULTING GROUP, LLC

**SDG No.:** Q1930

**Project:** Rotor Clip NJ WTD - 2025

**RunNo.:** LB135665

| Analyte | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|---------|-------|--------|------------|---------------|---------------------------|------------------|
|---------|-------|--------|------------|---------------|---------------------------|------------------|