



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Blank Summary

**Client:** VERINA CONSULTING GROUP, LLC

**SDG No.:** Q1930

**Project:** Rotor Clip NJ WTD - 2025

**RunNo.:** LB135627

| Analyte                                     | Units | Result   | Acceptance<br>Limits | Conc<br>Qual | MDL   | RDL | Analysis<br>Date |
|---------------------------------------------|-------|----------|----------------------|--------------|-------|-----|------------------|
| Sample ID: <b>ICB</b><br>Residual Chlorine  | mg/L  | < 0.0500 | 0.0500               | U            | 0.023 | 0.1 | 05/01/2025       |
| Sample ID: <b>CCB1</b><br>Residual Chlorine | mg/L  | < 0.0500 | 0.0500               | U            | 0.023 | 0.1 | 05/01/2025       |
| Sample ID: <b>CCB2</b><br>Residual Chlorine | mg/L  | < 0.0500 | 0.0500               | U            | 0.023 | 0.1 | 05/01/2025       |

### Initial and Continuing Calibration Blank Summary

**Client:** VERINA CONSULTING GROUP, LLC

**SDG No.:** Q1930

**Project:** Rotor Clip NJ WTD - 2025

**RunNo.:** LB135665

| Analyte                                | Units | Result   | Acceptance Limits | Conc Qual | MDL   | RDL | Analysis Date |
|----------------------------------------|-------|----------|-------------------|-----------|-------|-----|---------------|
| Sample ID: <b>ICB1</b><br>Ammonia as N | mg/L  | 0.034    | 0.0500            | J         | 0.030 | 0.1 | 05/05/2025    |
| Sample ID: <b>CCB1</b><br>Ammonia as N | mg/L  | 0.036    | 0.0500            | J         | 0.030 | 0.1 | 05/05/2025    |
| Sample ID: <b>CCB2</b><br>Ammonia as N | mg/L  | 0.039    | 0.0500            | J         | 0.030 | 0.1 | 05/05/2025    |
| Sample ID: <b>CCB3</b><br>Ammonia as N | mg/L  | 0.043    | 0.0500            | J         | 0.030 | 0.1 | 05/05/2025    |
| Sample ID: <b>CCB4</b><br>Ammonia as N | mg/L  | < 0.0500 | 0.0500            | U         | 0.030 | 0.1 | 05/05/2025    |