

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BLOCK #1 2. N/A

Start Digest Date: 05/05/2025 Time : 11:05 Temp : 96 °C

End Digest Date: 05/05/2025 Time : 14:10 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: *[Signature]*

Supervisor Signature: *[Signature]*

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6005
LFS-2	0.25	M6016
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6158
1:1 HCL	5.00	MP85156
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1CELL#50 96 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
05/05/25 13:10	<i>[Signature]</i> SPS.met.dig	<i>[Signature]</i> Metal Lab
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB167858BL	PBW858	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	1
PB167858BS	LCS858	<2	50	25	Colorless	Colorless	Clear	Clear	M6005,M6016	2
Q1930-01MS	WATER-TREATMENT-DISCHA RGEMS	<2	50	25	Colorless	Colorless	Clear	Clear	M6005,M6016	5
Q1930-01MSD	WATER-TREATMENT-DISCHA RGEMSD	<2	50	25	Colorless	Colorless	Clear	Clear	M6005,M6016	6
Q1930-01DUP	WATER-TREATMENT-DISCHA RGEDUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	4
Q1930-01	WATER-TREATMENT-DISCHA RGE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3
Q1944-02	CITY-WATER	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	7
Q1944-03	CHILLER-WATER	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	8
Q1945-01	GB1W	<2	50	25	Brown	light Brown	Clear	Clear	N/A	9
Q1956-07	FB-05022025	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	10