

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1930
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1930 **SAS No.:** Q1930
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	508	520	98	90 - 110	P	05/05/2025	14:18	LB135674
	Copper	499	510	98	90 - 110	P	05/05/2025	14:18	LB135674
	Nickel	490	530	92	90 - 110	P	05/05/2025	14:18	LB135674
	Zinc	969	1000	97	90 - 110	P	05/05/2025	14:18	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1930
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1930 **SAS No.:** Q1930
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	10.2	10.0	102	80 - 120	P	05/05/2025	14:36	LB135674
	Copper	22.4	20.0	112	80 - 120	P	05/05/2025	14:36	LB135674
	Nickel	39.5	40.0	99	80 - 120	P	05/05/2025	14:36	LB135674
	Zinc	41.2	40.0	103	80 - 120	P	05/05/2025	14:36	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1930

Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1930 **SAS No.:** Q1930

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	996	1000	100	90 - 110	P	05/05/2025	15:20	LB135674
	Copper	1240	1250	99	90 - 110	P	05/05/2025	15:20	LB135674
	Nickel	2450	2500	98	90 - 110	P	05/05/2025	15:20	LB135674
	Zinc	2480	2500	99	90 - 110	P	05/05/2025	15:20	LB135674
CCV02	Chromium	974	1000	97	90 - 110	P	05/05/2025	16:07	LB135674
	Copper	1230	1250	98	90 - 110	P	05/05/2025	16:07	LB135674
	Nickel	2420	2500	97	90 - 110	P	05/05/2025	16:07	LB135674
	Zinc	2430	2500	97	90 - 110	P	05/05/2025	16:07	LB135674
CCV03	Chromium	975	1000	98	90 - 110	P	05/05/2025	16:56	LB135674
	Copper	1260	1250	100	90 - 110	P	05/05/2025	16:56	LB135674
	Nickel	2420	2500	97	90 - 110	P	05/05/2025	16:56	LB135674
	Zinc	2460	2500	98	90 - 110	P	05/05/2025	16:56	LB135674
CCV04	Chromium	977	1000	98	90 - 110	P	05/05/2025	17:55	LB135674
	Copper	1220	1250	98	90 - 110	P	05/05/2025	17:55	LB135674
	Nickel	2420	2500	97	90 - 110	P	05/05/2025	17:55	LB135674
	Zinc	2420	2500	97	90 - 110	P	05/05/2025	17:55	LB135674
CCV05	Chromium	970	1000	97	90 - 110	P	05/05/2025	18:43	LB135674
	Copper	1230	1250	99	90 - 110	P	05/05/2025	18:43	LB135674
	Nickel	2430	2500	97	90 - 110	P	05/05/2025	18:43	LB135674
	Zinc	2470	2500	99	90 - 110	P	05/05/2025	18:43	LB135674
CCV06	Chromium	940	1000	94	90 - 110	P	05/05/2025	19:32	LB135674
	Copper	1180	1250	94	90 - 110	P	05/05/2025	19:32	LB135674
	Nickel	2360	2500	94	90 - 110	P	05/05/2025	19:32	LB135674
	Zinc	2380	2500	95	90 - 110	P	05/05/2025	19:32	LB135674
CCV07	Chromium	955	1000	96	90 - 110	P	05/05/2025	20:19	LB135674
	Copper	1170	1250	93	90 - 110	P	05/05/2025	20:19	LB135674
	Nickel	2360	2500	94	90 - 110	P	05/05/2025	20:19	LB135674
	Zinc	2350	2500	94	90 - 110	P	05/05/2025	20:19	LB135674
CCV08	Chromium	961	1000	96	90 - 110	P	05/05/2025	21:07	LB135674
	Copper	1190	1250	95	90 - 110	P	05/05/2025	21:07	LB135674
	Nickel	2380	2500	95	90 - 110	P	05/05/2025	21:07	LB135674
	Zinc	2350	2500	94	90 - 110	P	05/05/2025	21:07	LB135674
CCV09	Chromium	959	1000	96	90 - 110	P	05/05/2025	21:59	LB135674
	Copper	1180	1250	95	90 - 110	P	05/05/2025	21:59	LB135674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC SDG No.: Q1930

Contract: VERI01 Lab Code: CHEM Case No.: Q1930 SAS No.: Q1930

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV09	Nickel	2410	2500	96	90 - 110	P	05/05/2025	21:59	LB135674
	Zinc	2320	2500	93	90 - 110	P	05/05/2025	21:59	LB135674
CCV10	Chromium	967	1000	97	90 - 110	P	05/05/2025	22:33	LB135674
	Copper	1180	1250	95	90 - 110	P	05/05/2025	22:33	LB135674
	Nickel	2370	2500	95	90 - 110	P	05/05/2025	22:33	LB135674
	Zinc	2330	2500	93	90 - 110	P	05/05/2025	22:33	LB135674