



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1930

Contract: VERI01

Lab Code: CHEM

Case No.: Q1930

SAS No.: Q1930

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Chromium	10.3	10.0	103	40 - 160	P	05/05/2025	14:47	LB135674
	Copper	22.4	20.0	112	40 - 160	P	05/05/2025	14:47	LB135674
	Nickel	39.9	40.0	100	40 - 160	P	05/05/2025	14:47	LB135674
	Zinc	42.7	40.0	107	40 - 160	P	05/05/2025	14:47	LB135674