

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1930
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1930 **SAS No.:** Q1930
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	53.8	52.0	104	42	62	05/05/2025	14:52	LB135674
	Copper	12.1	2.0	605	-18	22	05/05/2025	14:52	LB135674
	Nickel	1.16	2.0	58	-38	42	05/05/2025	14:52	LB135674
	Zinc	3.23			-40	40	05/05/2025	14:52	LB135674
ICSAB01	Chromium	572	542	106	460	624	05/05/2025	15:07	LB135674
	Copper	520	511	102	434	588	05/05/2025	15:07	LB135674
	Nickel	1020	954	107	810	1100	05/05/2025	15:07	LB135674
	Zinc	1070	952	112	809	1095	05/05/2025	15:07	LB135674