

metals
- 5a -
MATRIX SPIKE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q1930
contract: VERI01 **lab code:** CHEM **case no.:** Q1930 **sas no.:** Q1930
matrix: Water **sample id:** Q1930-01 **client id:** WATER-TREATMENT-DISCHARGEMS
Percent Solids for Sample: NA **Spiked ID:** Q1930-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	345		156		200	94		P
Copper	ug/L	75 - 125	151		17.4		150	89		P
Nickel	ug/L	75 - 125	267		12.9	J	250	101		P
Zinc	ug/L	75 - 125	157		49.2		100	108		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q1930
contract: VERI01 **lab code:** CHEM **case no.:** Q1930 **sas no.:** Q1930
matrix: Water **sample id:** Q1930-01 **client id:** WATER-TREATMENT-DISCHARGEMSD
Percent Solids for Sample: NA **Spiked ID:** Q1930-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	341		156		200	93		P
Copper	ug/L	75 - 125	152		17.4		150	90		P
Nickel	ug/L	75 - 125	269		12.9	J	250	102		P
Zinc	ug/L	75 - 125	151		49.2		100	102		P