

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q1930
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1930 **SAS No.:** Q1930
Matrix: Water **Sample ID:** Q1930-01 **Client ID:** WATER-TREATMENT-DISCHARGEDUP
Percent Solids for Sample: NA **Duplicate ID** Q1930-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	156		159		2		P
Copper	ug/L	20	17.4		17.4		0		P
Nickel	ug/L	20	12.9	J	12.8	J	1		P
Zinc	ug/L	20	49.2		49.8		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q1930
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1930 **SAS No.:** Q1930
Matrix: Water **Sample ID:** Q1930-01MS **Client ID:** WATER-TREATMENT-DISCHARGEMSD
Percent Solids for Sample: NA **Duplicate ID** Q1930-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	345		341		1		P
Copper	ug/L	20	151		152		1		P
Nickel	ug/L	20	267		269		1		P
Zinc	ug/L	20	157		151		4		P