

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1930

Contract: VERI01

Lab Code: CHEM

Case No.: Q1930

SAS No.: Q1930

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167858BS							
Chromium	ug/L	200	183		92	80 - 120	P
Copper	ug/L	150	140		93	80 - 120	P
Nickel	ug/L	250	228		91	80 - 120	P
Zinc	ug/L	100	91.9		92	80 - 120	P