



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Tully Environmental, Inc

SDG No.: Q1934

Project: Transfer Station-SPDES

RunNo.: LB135665

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	0.034	0.0500	J	0.030	0.1	05/05/2025
Sample ID: CCB1 Ammonia as N	mg/L	0.036	0.0500	J	0.030	0.1	05/05/2025
Sample ID: CCB2 Ammonia as N	mg/L	0.039	0.0500	J	0.030	0.1	05/05/2025
Sample ID: CCB3 Ammonia as N	mg/L	0.043	0.0500	J	0.030	0.1	05/05/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	05/05/2025