

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1934

Project: Transfer Station-SPDES

Run No.: LB135665

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167844BS							
Ammonia as N	mg/L	1	0.96		96	1	90-110	05/05/2025