

Continuing Calibration Report for SequenceID : FE050525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE053640.D

Aliphatic C9-C12	9038765.000	60.000	3.112	6.755	150646.083	138562.354	-8.721
Aliphatic C12-C16	6131080.000	40.000	6.756	10.204	153277.000	141806.081	-8.089
Aliphatic C16-C21	9075695.000	60.000	10.205	13.579	151261.583	145363.765	-4.057
Aliphatic C21-C28	11352728.000	80.000	13.580	17.249	141909.100	142174.626	0.187
Aliphatic C28-C40	14676919.000	120.000	17.250	22.140	122307.658	129049.185	5.224
Aliphatic EPH	50275187.000	360.000	3.112	22.140	139653.297	137688.007	-1.427