

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1936 SAS No.: Q1936 SDG NO.: Q1936

Continuing Calib Date: 05/05/2025 Initial Calibration Date(s): 05/02/2025 05/02/2025

Continuing Calib Time: 10:58 Initial Calibration Time(s): 09:45 18:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.58	5.58	5.48	5.68	0.00
Aroclor-1254-2	(2)	5.73	5.73	5.63	5.83	0.00
Aroclor-1254-3	(3)	6.14	6.14	6.04	6.24	0.00
Aroclor-1254-4	(4)	6.37	6.37	6.27	6.47	0.01
Aroclor-1254-5	(5)	6.79	6.79	6.69	6.89	0.01
Tetrachloro-m-xylene		3.69	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1936 SAS No.: Q1936 SDG NO.: Q1936

Continuing Calib Date: 05/05/2025 Initial Calibration Date(s): 05/02/2025 05/02/2025

Continuing Calib Time: 10:58 Initial Calibration Time(s): 09:45 18:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.56	5.56	5.46	5.66	0.00
Aroclor-1254-2	(2)	5.71	5.71	5.61	5.81	0.00
Aroclor-1254-3	(3)	6.11	6.11	6.01	6.21	0.00
Aroclor-1254-4	(4)	6.34	6.34	6.24	6.44	0.00
Aroclor-1254-5	(5)	6.75	6.76	6.66	6.86	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.67	8.68	8.58	8.78	0.01