

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1936 SAS No.: Q1936 SDG NO.: Q1936

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL01 Date Analyzed: 05/05/2025

Lab Sample No.: AR1660CCC500 Data File : PO110926.D Time Analyzed: 10:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.780	4.673	4.873	449.500	500.000	-10.1
Aroclor-1016-2	4.800	4.692	4.892	462.350	500.000	-7.5
Aroclor-1016-3	4.856	4.748	4.948	431.450	500.000	-13.7
Aroclor-1016-4	4.976	4.868	5.068	459.620	500.000	-8.1
Aroclor-1016-5	5.234	5.125	5.325	449.770	500.000	-10.0
Aroclor-1260-1	6.275	6.165	6.365	441.930	500.000	-11.6
Aroclor-1260-2	6.463	6.355	6.555	440.130	500.000	-12.0
Aroclor-1260-3	6.831	6.722	6.922	464.700	500.000	-7.1
Aroclor-1260-4	7.091	6.982	7.182	455.430	500.000	-8.9
Aroclor-1260-5	7.334	7.225	7.425	459.130	500.000	-8.2
Decachlorobiphenyl	8.736	8.625	8.825	45.470	50.000	-9.1
Tetrachloro-m-xylene	3.689	3.583	3.783	46.640	50.000	-6.7

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1936 SAS No.: Q1936 SDG NO.: Q1936

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL01 Date Analyzed: 05/05/2025

Lab Sample No.: AR1660CCC500 Data File : PO110926.D Time Analyzed: 10:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.760	4.660	4.860	485.320	500.000	-2.9
Aroclor-1016-2	4.779	4.679	4.879	485.430	500.000	-2.9
Aroclor-1016-3	4.954	4.854	5.054	491.110	500.000	-1.8
Aroclor-1016-4	4.996	4.896	5.096	494.570	500.000	-1.1
Aroclor-1016-5	5.208	5.109	5.309	483.070	500.000	-3.4
Aroclor-1260-1	6.240	6.139	6.339	480.420	500.000	-3.9
Aroclor-1260-2	6.428	6.327	6.527	479.660	500.000	-4.1
Aroclor-1260-3	6.580	6.480	6.680	493.460	500.000	-1.3
Aroclor-1260-4	7.052	6.950	7.150	496.890	500.000	-0.6
Aroclor-1260-5	7.292	7.192	7.392	505.320	500.000	1.1
Decachlorobiphenyl	8.677	8.576	8.776	52.350	50.000	4.7
Tetrachloro-m-xylene	3.680	3.580	3.780	50.010	50.000	0.0