

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050225\
 Data File : P0110922.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2025 20:54
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO050225AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 23:35:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 02 23:33:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.684	3.680	281.7E6	249.0E6	50.508	53.918
2) SA Decachlor...	8.728	8.676	291.0E6	86640442	47.301	50.921
Target Compounds						
21) L5 AR-1248-1	4.775	4.759	65440838	50906724	498.171	528.215
22) L5 AR-1248-2	5.013	4.996	87125763	69718152	497.201	515.965
23) L5 AR-1248-3	5.228	5.038	109.3E6	74471131	495.964	513.088
24) L5 AR-1248-4	5.582	5.208	157.6E6	87700628	480.768	523.535
25) L5 AR-1248-5	5.623	5.600	113.6E6	86365939	486.636	514.455



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050225\
Data File : PO110922.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2025 20:54
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO050225AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 23:35:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050225.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 02 23:33:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.684 min
Delta R.T.: 0.001 min
Response: 281652415
Conc: 50.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO050225AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.680 min
Delta R.T.: 0.000 min
Response: 248964908
Conc: 53.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.728 min
Delta R.T.: 0.002 min
Response: 291010640
Conc: 47.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.676 min
Delta R.T.: 0.000 min
Response: 86640442
Conc: 50.92 ng/ml



#21 AR-1248-1


R.T.: 4.775 min
Delta R.T.: 0.001 min
Response: 65440838
Conc: 498.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO050225AR1248




#21 AR-1248-1

R.T.: 4.759 min
Delta R.T.: 0.000 min
Response: 50906724
Conc: 528.22 ng/ml




#22 AR-1248-2

R.T.: 5.013 min
Delta R.T.: 0.002 min
Response: 87125763
Conc: 497.20 ng/ml



#22 AR-1248-2


R.T.: 4.996 min
Delta R.T.: 0.000 min
Response: 69718152
Conc: 515.97 ng/ml



#23 AR-1248-3


R.T.: 5.228 min
Delta R.T.: 0.002 min
Response: 109344102
Conc: 495.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO050225AR1248




#23 AR-1248-3

R.T.: 5.038 min
Delta R.T.: 0.000 min
Response: 74471131
Conc: 513.09 ng/ml




#24 AR-1248-4

R.T.: 5.582 min
Delta R.T.: 0.002 min
Response: 157554622
Conc: 480.77 ng/ml



#24 AR-1248-4


R.T.: 5.208 min
Delta R.T.: 0.000 min
Response: 87700628
Conc: 523.54 ng/ml



#25 AR-1248-5

R.T.: 5.623 min
Delta R.T.: 0.001 min
Response: 113604147
Conc: 486.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO050225AR1248



#25 AR-1248-5

R.T.: 5.600 min
Delta R.T.: 0.000 min
Response: 86365939
Conc: 514.45 ng/ml