

Surrogate Summary

SDG No.: Q1936

Client: Saxton Falls Sand and Gravel Co. Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO110891.D	PIBLK-PO110891.D	Tetrachloro-m-xylene	1	20	16.1	81		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.1	91		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	15.2	76		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	18.1	90		70 (60)	130 (140)
I.BLK-PO110930.D	PIBLK-PO110930.D	Tetrachloro-m-xylene	1	20	14.9	75		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	16.1	81		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	15.5	78		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	18.0	90		70 (60)	130 (140)
PB167853BL	PB167853BL	Tetrachloro-m-xylene	1	20	18.3	91		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	18.8	94		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.2	91		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.6	103		30 (32)	150 (175)
PB167853BS	PB167853BS	Tetrachloro-m-xylene	1	20	17.0	85		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.0	95		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.5	87		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.6	103		30 (32)	150 (175)
Q1936-01	SMALL-PILE-A	Tetrachloro-m-xylene	1	20	14.2	71		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	13.4	67		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	14.6	73		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	14.8	74		30 (32)	150 (175)
Q1936-01MS	SMALL-PILE-AMS	Tetrachloro-m-xylene	1	20	17.0	85		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	16.6	83		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.9	89		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.7	93		30 (32)	150 (175)
Q1936-01MSD	SMALL-PILE-AMSD	Tetrachloro-m-xylene	1	20	17.1	85		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	16.6	83		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.0	90		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.0	95		30 (32)	150 (175)
Q1936-03	SMALL-PILE-B	Tetrachloro-m-xylene	1	20	18.0	90		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	17.9	89		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.7	93		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.1	106		30 (32)	150 (175)
Q1936-05	SMALL-PILE-C	Tetrachloro-m-xylene	1	20	15.0	75		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	14.6	73		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	15.9	79		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	17.2	86		30 (32)	150 (175)
Q1936-07	SMALL-PILE-D	Tetrachloro-m-xylene	1	20	12.2	61		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	12.2	61		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	12.0	60		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	14.0	70		30 (32)	150 (175)
I.BLK-PO110945.D	PIBLK-PO110945.D	Tetrachloro-m-xylene	1	20	15.4	77		70 (60)	130 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q1936

Client: Saxton Falls Sand and Gravel Co. Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO110945.D	PIBLK-PO110945.D	Decachlorobiphenyl	1	20	15.4	77		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	16.2	81		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.9	90		70 (60)	130 (140)