

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1936

Client: Saxton Falls Sand and Gravel Co. Inc.

Analytical Method: **8081B**

Datafile : PL095532.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167854BS	alpha-BHC	16.65	20.4	ug/kg	123				40 (84)	140 (123)	
	beta-BHC	16.65	18.3	ug/kg	110				40 (82)	140 (123)	
	delta-BHC	16.65	20.0	ug/kg	120				40 (83)	140 (126)	
	gamma-BHC (Lindane)	16.65	19.5	ug/kg	117				40 (83)	140 (125)	
	Heptachlor	16.65	18.9	ug/kg	114				40 (83)	140 (122)	
	Aldrin	16.65	20.2	ug/kg	121				40 (82)	140 (124)	
	Heptachlor epoxide	16.65	19.5	ug/kg	117				40 (83)	140 (120)	
	Endosulfan I	16.65	19.6	ug/kg	118				40 (81)	140 (124)	
	Dieldrin	16.65	19.6	ug/kg	118				40 (85)	140 (121)	
	4,4'-DDE	16.65	18.2	ug/kg	109				40 (81)	140 (123)	
	Endrin	16.65	17.0	ug/kg	102				40 (76)	140 (130)	
	Endosulfan II	16.65	19.3	ug/kg	116				40 (80)	140 (125)	
	4,4'-DDD	16.65	20.4	ug/kg	123				40 (80)	140 (131)	
	Endosulfan sulfate	16.65	19.3	ug/kg	116				40 (81)	140 (122)	
	4,4'-DDT	16.65	17.3	ug/kg	104				40 (70)	140 (129)	
	Methoxychlor	16.65	17.2	ug/kg	103				40 (60)	140 (119)	
	Endrin ketone	16.65	21.4	ug/kg	129				40 (77)	140 (132)	
	Endrin aldehyde	16.65	19.9	ug/kg	120				40 (79)	140 (124)	
	alpha-Chlordane	16.65	19.2	ug/kg	115				40 (84)	140 (120)	
	gamma-Chlordane	16.65	19.4	ug/kg	117				40 (83)	140 (122)	