

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1936 SAS No.: Q1936 SDG NO.: Q1936

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/05/2025

Lab File ID: BM050096.D Time Analyzed: 12:18

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	253723	7.745	967980	10.54	663376	14.40
UPPER LIMIT	507446	8.245	1935960	11.039	1326750	14.898
LOWER LIMIT	126862	7.245	483990	10.039	331688	13.898
EPA SAMPLE NO.						
01 SMALL-PILE-B	282446	7.75	1024680	10.54	673586	14.39
02 LARGE-PILE-CMS	289071	7.75	1033170	10.54	655158	14.39
03 LARGE-PILE-CMSD	294539	7.75	1047790	10.54	677703	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.