

Hit Summary Sheet
SW-846

SDG No.: Q1936
Client: Saxton Falls Sand and Gravel Co. Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: Q1936-04	SMALL-PILE-B SMALL-PILE-B	SOIL	Methylene Chloride	3.80	J	3.60	10.1	ug/Kg
			Total Voc :	3.80				
			Total Concentration:	3.80				
Client ID: Q1936-06	SMALL-PILE-C SMALL-PILE-C	SOIL	Methylene Chloride	4.00	J	2.60	7.40	ug/Kg
			Total Voc :	4.00				
			Total Concentration:	4.00				
Client ID: Q1936-08	SMALL-PILE-D SMALL-PILE-D	SOIL	Methylene Chloride	3.70	J	3.20	8.90	ug/Kg
			Total Voc :	3.70				
			Total Concentration:	3.70				

A

B

C

D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-A	SDG No.:	Q1936
Lab Sample ID:	Q1936-02	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	91
Sample Wt/Vol:	6.13 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022142.D	1		05/02/25 20:19	VY050225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.00	U	1.00	4.50	ug/Kg
74-87-3	Chloromethane	1.00	U	1.00	4.50	ug/Kg
75-01-4	Vinyl Chloride	0.71	U	0.71	4.50	ug/Kg
74-83-9	Bromomethane	0.96	U	0.96	4.50	ug/Kg
75-00-3	Chloroethane	1.10	U	1.10	4.50	ug/Kg
75-69-4	Trichlorofluoromethane	1.10	U	1.10	4.50	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.95	U	0.95	4.50	ug/Kg
75-35-4	1,1-Dichloroethene	0.90	U	0.90	4.50	ug/Kg
67-64-1	Acetone	4.20	U	4.20	22.4	ug/Kg
75-15-0	Carbon Disulfide	0.95	U	0.95	4.50	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.65	U	0.65	4.50	ug/Kg
79-20-9	Methyl Acetate	1.40	U	1.40	4.50	ug/Kg
75-09-2	Methylene Chloride	3.20	U	3.20	9.00	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.77	U	0.77	4.50	ug/Kg
75-34-3	1,1-Dichloroethane	0.72	U	0.72	4.50	ug/Kg
110-82-7	Cyclohexane	0.71	U	0.71	4.50	ug/Kg
78-93-3	2-Butanone	5.90	U	5.90	22.4	ug/Kg
56-23-5	Carbon Tetrachloride	0.87	U	0.87	4.50	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.67	U	0.67	4.50	ug/Kg
74-97-5	Bromochloromethane	1.00	U	1.00	4.50	ug/Kg
67-66-3	Chloroform	0.75	U	0.75	4.50	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.83	U	0.83	4.50	ug/Kg
108-87-2	Methylcyclohexane	0.82	U	0.82	4.50	ug/Kg
71-43-2	Benzene	0.71	U	0.71	4.50	ug/Kg
107-06-2	1,2-Dichloroethane	0.71	U	0.71	4.50	ug/Kg
79-01-6	Trichloroethene	0.73	U	0.73	4.50	ug/Kg
78-87-5	1,2-Dichloropropane	0.82	U	0.82	4.50	ug/Kg
75-27-4	Bromodichloromethane	0.70	U	0.70	4.50	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.20	U	3.20	22.4	ug/Kg
108-88-3	Toluene	0.70	U	0.70	4.50	ug/Kg

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-A	SDG No.:	Q1936
Lab Sample ID:	Q1936-02	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	91
Sample Wt/Vol:	6.13 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022142.D	1		05/02/25 20:19	VY050225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.58	U	0.58	4.50	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.56	U	0.56	4.50	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.82	U	0.82	4.50	ug/Kg
591-78-6	2-Hexanone	3.30	U	3.30	22.4	ug/Kg
124-48-1	Dibromochloromethane	0.78	U	0.78	4.50	ug/Kg
106-93-4	1,2-Dibromoethane	0.79	U	0.79	4.50	ug/Kg
127-18-4	Tetrachloroethene	0.94	U	0.94	4.50	ug/Kg
108-90-7	Chlorobenzene	0.82	U	0.82	4.50	ug/Kg
100-41-4	Ethyl Benzene	0.60	U	0.60	4.50	ug/Kg
179601-23-1	m/p-Xylenes	1.10	U	1.10	9.00	ug/Kg
95-47-6	o-Xylene	0.73	U	0.73	4.50	ug/Kg
100-42-5	Styrene	0.64	U	0.64	4.50	ug/Kg
75-25-2	Bromoform	0.77	U	0.77	4.50	ug/Kg
98-82-8	Isopropylbenzene	0.70	U	0.70	4.50	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.10	U	1.10	4.50	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.50	U	1.50	4.50	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.40	U	1.40	4.50	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.30	U	1.30	4.50	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.60	U	1.60	4.50	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.70	U	2.70	4.50	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.90	U	2.90	4.50	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.7		70 (63) - 130 (155)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	49.1		70 (70) - 130 (134)	98%	SPK: 50
2037-26-5	Toluene-d8	48.0		70 (74) - 130 (123)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	40.1		70 (38) - 130 (136)	80%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	328000	7.707			
540-36-3	1,4-Difluorobenzene	626000	8.609			
3114-55-4	Chlorobenzene-d5	546000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	207000	13.346			

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-A	SDG No.:	Q1936
Lab Sample ID:	Q1936-02	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	91
Sample Wt/Vol:	6.13 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022142.D	1		05/02/25 20:19	VY050225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-B	SDG No.:	Q1936
Lab Sample ID:	Q1936-04	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	93.7
Sample Wt/Vol:	5.27 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022159.D	1		05/05/25 15:08	VY050525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.20	U	1.20	5.10	ug/Kg
74-87-3	Chloromethane	1.20	U	1.20	5.10	ug/Kg
75-01-4	Vinyl Chloride	0.80	U	0.80	5.10	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.10	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.10	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	5.10	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.10	ug/Kg
75-35-4	1,1-Dichloroethene	1.00	U	1.00	5.10	ug/Kg
67-64-1	Acetone	4.80	U	4.80	25.3	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.10	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.74	U	0.74	5.10	ug/Kg
79-20-9	Methyl Acetate	1.60	U	1.60	5.10	ug/Kg
75-09-2	Methylene Chloride	3.80	J	3.60	10.1	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.87	U	0.87	5.10	ug/Kg
75-34-3	1,1-Dichloroethane	0.81	U	0.81	5.10	ug/Kg
110-82-7	Cyclohexane	0.80	U	0.80	5.10	ug/Kg
78-93-3	2-Butanone	6.60	U	6.60	25.3	ug/Kg
56-23-5	Carbon Tetrachloride	0.98	U	0.98	5.10	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.76	U	0.76	5.10	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.10	ug/Kg
67-66-3	Chloroform	0.85	U	0.85	5.10	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.94	U	0.94	5.10	ug/Kg
108-87-2	Methylcyclohexane	0.92	U	0.92	5.10	ug/Kg
71-43-2	Benzene	0.80	U	0.80	5.10	ug/Kg
107-06-2	1,2-Dichloroethane	0.80	U	0.80	5.10	ug/Kg
79-01-6	Trichloroethene	0.82	U	0.82	5.10	ug/Kg
78-87-5	1,2-Dichloropropane	0.92	U	0.92	5.10	ug/Kg
75-27-4	Bromodichloromethane	0.79	U	0.79	5.10	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.60	U	3.60	25.3	ug/Kg
108-88-3	Toluene	0.79	U	0.79	5.10	ug/Kg

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-B	SDG No.:	Q1936
Lab Sample ID:	Q1936-04	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	93.7
Sample Wt/Vol:	5.27 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022159.D	1		05/05/25 15:08	VY050525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.66	U	0.66	5.10	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.63	U	0.63	5.10	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.93	U	0.93	5.10	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	25.3	ug/Kg
124-48-1	Dibromochloromethane	0.88	U	0.88	5.10	ug/Kg
106-93-4	1,2-Dibromoethane	0.89	U	0.89	5.10	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.10	ug/Kg
108-90-7	Chlorobenzene	0.92	U	0.92	5.10	ug/Kg
100-41-4	Ethyl Benzene	0.68	U	0.68	5.10	ug/Kg
179601-23-1	m/p-Xylenes	1.30	U	1.30	10.1	ug/Kg
95-47-6	o-Xylene	0.83	U	0.83	5.10	ug/Kg
100-42-5	Styrene	0.72	U	0.72	5.10	ug/Kg
75-25-2	Bromoform	0.87	U	0.87	5.10	ug/Kg
98-82-8	Isopropylbenzene	0.79	U	0.79	5.10	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	5.10	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	5.10	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.60	U	1.60	5.10	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.10	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.90	U	1.90	5.10	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.00	U	3.00	5.10	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.20	U	3.20	5.10	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.1		70 (63) - 130 (155)	110%	SPK: 50
1868-53-7	Dibromofluoromethane	50.6		70 (70) - 130 (134)	101%	SPK: 50
2037-26-5	Toluene-d8	49.1		70 (74) - 130 (123)	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	38.5		70 (38) - 130 (136)	77%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	228000	7.707			
540-36-3	1,4-Difluorobenzene	442000	8.615			
3114-55-4	Chlorobenzene-d5	390000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	141000	13.346			

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-B	SDG No.:	Q1936
Lab Sample ID:	Q1936-04	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	93.7
Sample Wt/Vol:	5.27 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022159.D	1		05/05/25 15:08	VY050525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-C	SDG No.:	Q1936
Lab Sample ID:	Q1936-06	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	96.2
Sample Wt/Vol:	7.03 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022160.D	1		05/05/25 15:31	VY050525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.84	U	0.84	3.70	ug/Kg
74-87-3	Chloromethane	0.84	U	0.84	3.70	ug/Kg
75-01-4	Vinyl Chloride	0.58	U	0.58	3.70	ug/Kg
74-83-9	Bromomethane	0.79	U	0.79	3.70	ug/Kg
75-00-3	Chloroethane	0.93	U	0.93	3.70	ug/Kg
75-69-4	Trichlorofluoromethane	0.89	U	0.89	3.70	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.78	U	0.78	3.70	ug/Kg
75-35-4	1,1-Dichloroethene	0.74	U	0.74	3.70	ug/Kg
67-64-1	Acetone	3.50	U	3.50	18.5	ug/Kg
75-15-0	Carbon Disulfide	0.78	U	0.78	3.70	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.54	U	0.54	3.70	ug/Kg
79-20-9	Methyl Acetate	1.10	U	1.10	3.70	ug/Kg
75-09-2	Methylene Chloride	4.00	J	2.60	7.40	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.64	U	0.64	3.70	ug/Kg
75-34-3	1,1-Dichloroethane	0.59	U	0.59	3.70	ug/Kg
110-82-7	Cyclohexane	0.58	U	0.58	3.70	ug/Kg
78-93-3	2-Butanone	4.80	U	4.80	18.5	ug/Kg
56-23-5	Carbon Tetrachloride	0.72	U	0.72	3.70	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.55	U	0.55	3.70	ug/Kg
74-97-5	Bromochloromethane	0.85	U	0.85	3.70	ug/Kg
67-66-3	Chloroform	0.62	U	0.62	3.70	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.69	U	0.69	3.70	ug/Kg
108-87-2	Methylcyclohexane	0.67	U	0.67	3.70	ug/Kg
71-43-2	Benzene	0.58	U	0.58	3.70	ug/Kg
107-06-2	1,2-Dichloroethane	0.58	U	0.58	3.70	ug/Kg
79-01-6	Trichloroethene	0.60	U	0.60	3.70	ug/Kg
78-87-5	1,2-Dichloropropane	0.67	U	0.67	3.70	ug/Kg
75-27-4	Bromodichloromethane	0.58	U	0.58	3.70	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2.60	U	2.60	18.5	ug/Kg
108-88-3	Toluene	0.58	U	0.58	3.70	ug/Kg

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-C	SDG No.:	Q1936
Lab Sample ID:	Q1936-06	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	96.2
Sample Wt/Vol:	7.03 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022160.D	1		05/05/25 15:31	VY050525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.48	U	0.48	3.70	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.46	U	0.46	3.70	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.68	U	0.68	3.70	ug/Kg
591-78-6	2-Hexanone	2.70	U	2.70	18.5	ug/Kg
124-48-1	Dibromochloromethane	0.64	U	0.64	3.70	ug/Kg
106-93-4	1,2-Dibromoethane	0.65	U	0.65	3.70	ug/Kg
127-18-4	Tetrachloroethene	0.78	U	0.78	3.70	ug/Kg
108-90-7	Chlorobenzene	0.67	U	0.67	3.70	ug/Kg
100-41-4	Ethyl Benzene	0.50	U	0.50	3.70	ug/Kg
179601-23-1	m/p-Xylenes	0.92	U	0.92	7.40	ug/Kg
95-47-6	o-Xylene	0.61	U	0.61	3.70	ug/Kg
100-42-5	Styrene	0.52	U	0.52	3.70	ug/Kg
75-25-2	Bromoform	0.64	U	0.64	3.70	ug/Kg
98-82-8	Isopropylbenzene	0.58	U	0.58	3.70	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.89	U	0.89	3.70	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.30	U	1.30	3.70	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.20	U	1.20	3.70	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.10	U	1.10	3.70	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.40	U	1.40	3.70	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.20	U	2.20	3.70	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.40	U	2.40	3.70	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	56.5		70 (63) - 130 (155)	113%	SPK: 50
1868-53-7	Dibromofluoromethane	50.9		70 (70) - 130 (134)	102%	SPK: 50
2037-26-5	Toluene-d8	49.1		70 (74) - 130 (123)	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	39.0		70 (38) - 130 (136)	78%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	226000	7.713			
540-36-3	1,4-Difluorobenzene	438000	8.616			
3114-55-4	Chlorobenzene-d5	386000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	139000	13.346			

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-C	SDG No.:	Q1936
Lab Sample ID:	Q1936-06	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	96.2
Sample Wt/Vol:	7.03 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022160.D	1		05/05/25 15:31	VY050525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-D	SDG No.:	Q1936
Lab Sample ID:	Q1936-08	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	94.1
Sample Wt/Vol:	5.95 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022161.D	1		05/05/25 15:55	VY050525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.00	U	1.00	4.50	ug/Kg
74-87-3	Chloromethane	1.00	U	1.00	4.50	ug/Kg
75-01-4	Vinyl Chloride	0.71	U	0.71	4.50	ug/Kg
74-83-9	Bromomethane	0.96	U	0.96	4.50	ug/Kg
75-00-3	Chloroethane	1.10	U	1.10	4.50	ug/Kg
75-69-4	Trichlorofluoromethane	1.10	U	1.10	4.50	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.95	U	0.95	4.50	ug/Kg
75-35-4	1,1-Dichloroethene	0.89	U	0.89	4.50	ug/Kg
67-64-1	Acetone	4.20	U	4.20	22.3	ug/Kg
75-15-0	Carbon Disulfide	0.95	U	0.95	4.50	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.65	U	0.65	4.50	ug/Kg
79-20-9	Methyl Acetate	1.40	U	1.40	4.50	ug/Kg
75-09-2	Methylene Chloride	3.70	J	3.20	8.90	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.77	U	0.77	4.50	ug/Kg
75-34-3	1,1-Dichloroethane	0.71	U	0.71	4.50	ug/Kg
110-82-7	Cyclohexane	0.71	U	0.71	4.50	ug/Kg
78-93-3	2-Butanone	5.80	U	5.80	22.3	ug/Kg
56-23-5	Carbon Tetrachloride	0.87	U	0.87	4.50	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.67	U	0.67	4.50	ug/Kg
74-97-5	Bromochloromethane	1.00	U	1.00	4.50	ug/Kg
67-66-3	Chloroform	0.75	U	0.75	4.50	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.83	U	0.83	4.50	ug/Kg
108-87-2	Methylcyclohexane	0.81	U	0.81	4.50	ug/Kg
71-43-2	Benzene	0.71	U	0.71	4.50	ug/Kg
107-06-2	1,2-Dichloroethane	0.71	U	0.71	4.50	ug/Kg
79-01-6	Trichloroethene	0.72	U	0.72	4.50	ug/Kg
78-87-5	1,2-Dichloropropane	0.81	U	0.81	4.50	ug/Kg
75-27-4	Bromodichloromethane	0.70	U	0.70	4.50	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.20	U	3.20	22.3	ug/Kg
108-88-3	Toluene	0.70	U	0.70	4.50	ug/Kg

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.	Date Collected:	05/01/25
Project:	Stan Hope	Date Received:	05/01/25
Client Sample ID:	SMALL-PILE-D	SDG No.:	Q1936
Lab Sample ID:	Q1936-08	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	94.1
Sample Wt/Vol:	5.95 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022161.D	1		05/05/25 15:55	VY050525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.58	U	0.58	4.50	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.55	U	0.55	4.50	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.82	U	0.82	4.50	ug/Kg
591-78-6	2-Hexanone	3.30	U	3.30	22.3	ug/Kg
124-48-1	Dibromochloromethane	0.78	U	0.78	4.50	ug/Kg
106-93-4	1,2-Dibromoethane	0.79	U	0.79	4.50	ug/Kg
127-18-4	Tetrachloroethene	0.94	U	0.94	4.50	ug/Kg
108-90-7	Chlorobenzene	0.81	U	0.81	4.50	ug/Kg
100-41-4	Ethyl Benzene	0.60	U	0.60	4.50	ug/Kg
179601-23-1	m/p-Xylenes	1.10	U	1.10	8.90	ug/Kg
95-47-6	o-Xylene	0.73	U	0.73	4.50	ug/Kg
100-42-5	Styrene	0.63	U	0.63	4.50	ug/Kg
75-25-2	Bromoform	0.77	U	0.77	4.50	ug/Kg
98-82-8	Isopropylbenzene	0.70	U	0.70	4.50	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.10	U	1.10	4.50	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.50	U	1.50	4.50	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.40	U	1.40	4.50	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.30	U	1.30	4.50	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.60	U	1.60	4.50	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.70	U	2.70	4.50	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.80	U	2.80	4.50	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	57.6		70 (63) - 130 (155)	115%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		70 (70) - 130 (134)	102%	SPK: 50
2037-26-5	Toluene-d8	49.4		70 (74) - 130 (123)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	40.0		70 (38) - 130 (136)	80%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	221000	7.713			
540-36-3	1,4-Difluorobenzene	435000	8.616			
3114-55-4	Chlorobenzene-d5	389000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	142000	13.346			

Report of Analysis

Client:	Saxton Falls Sand and Gravel Co. Inc.		Date Collected:	05/01/25	
Project:	Stan Hope		Date Received:	05/01/25	
Client Sample ID:	SMALL-PILE-D		SDG No.:	Q1936	
Lab Sample ID:	Q1936-08		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	94.1	
Sample Wt/Vol:	5.95	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022161.D	1		05/05/25 15:55	VY050525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q1936	OrderDate:	5/1/2025 2:00:17 PM
Client:	Saxton Falls Sand and Gravel Co. Inc.	Project:	Stan Hope
Contact:	Rich Schindelar	Location:	L51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1936-02	SMALL-PILE-A	SOIL	VOC-TCLVOA-10	8260D	05/01/25		05/02/25	05/01/25
Q1936-04	SMALL-PILE-B	SOIL	VOC-TCLVOA-10	8260D	05/01/25		05/05/25	05/01/25
Q1936-06	SMALL-PILE-C	SOIL	VOC-TCLVOA-10	8260D	05/01/25		05/05/25	05/01/25
Q1936-08	SMALL-PILE-D	SOIL	VOC-TCLVOA-10	8260D	05/01/25		05/05/25	05/01/25