

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1937 SAS No.: Q1937 SDG NO.: Q1937

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP071766.D Time Analyzed: 10:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.666	5.571	5.771	495.490	500.000	-0.9
Aroclor-1016-2	5.687	5.593	5.793	504.970	500.000	1.0
Aroclor-1016-3	5.749	5.654	5.854	504.560	500.000	0.9
Aroclor-1016-4	5.847	5.752	5.952	506.620	500.000	1.3
Aroclor-1016-5	6.139	6.045	6.245	501.100	500.000	0.2
Aroclor-1260-1	7.257	7.165	7.365	479.580	500.000	-4.1
Aroclor-1260-2	7.511	7.418	7.618	491.060	500.000	-1.8
Aroclor-1260-3	7.869	7.777	7.977	509.620	500.000	1.9
Aroclor-1260-4	8.093	8.002	8.202	499.990	500.000	0.0
Aroclor-1260-5	8.413	8.322	8.522	520.850	500.000	4.2
Decachlorobiphenyl	10.224	10.137	10.337	52.260	50.000	4.5
Tetrachloro-m-xylene	4.512	4.417	4.617	51.370	50.000	2.7

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1937 SAS No.: Q1937 SDG NO.: Q1937

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 05/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP071766.D Time Analyzed: 10:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.892	4.798	4.998	470.630	500.000	-5.9
Aroclor-1016-2	4.910	4.816	5.016	493.690	500.000	-1.3
Aroclor-1016-3	5.087	4.993	5.193	501.580	500.000	0.3
Aroclor-1016-4	5.128	5.036	5.236	501.560	500.000	0.3
Aroclor-1016-5	5.343	5.250	5.450	496.800	500.000	-0.6
Aroclor-1260-1	6.376	6.285	6.485	497.390	500.000	-0.5
Aroclor-1260-2	6.564	6.473	6.673	502.000	500.000	0.4
Aroclor-1260-3	6.717	6.627	6.827	489.370	500.000	-2.1
Aroclor-1260-4	7.188	7.098	7.298	493.060	500.000	-1.4
Aroclor-1260-5	7.429	7.339	7.539	476.020	500.000	-4.8
Decachlorobiphenyl	8.835	8.749	8.949	50.840	50.000	1.7
Tetrachloro-m-xylene	3.807	3.712	3.912	50.490	50.000	1.0