

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: SAXT01
ProjectID: Stan Hope
Lab Code: CHEM Case No.: Q1937 SAS No.: Q1937 SDG No.: Q1937
DataFile: FG015784.D Analyst Name: YP\AJ Analyst Date: 05-07-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	94713719	111428	121213	8.073