

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: SAXT01  
 Lab Code: CHEM Case No.: Q1937 SAS No.: Q1937 SDG No.: Q1937  
 Instrument ID: BNA\_P Calibration Date(s): 04/14/2025 04/14/2025  
 Calibration Time(s): 11:06 17:13

LAB FILE ID:								
RRF2.5 = BP024275.D			RRF005 = BP024276.D			RRF010 = BP024277.D		
RRF020 = BP024278.D			RRF040 = BP024279.D			RRF050 = BP024280.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.123	1.145	1.262	1.254	1.210	1.210	4.9
Benzaldehyde		0.943	0.934	0.942	0.836	0.732	0.819	15.7
Phenol-d6		1.463	1.552	1.718	1.748	1.697	1.656	6.7
Phenol		1.475	1.565	1.720	1.750	1.698	1.661	6.5
bis(2-Chloroethyl)ether		1.278	1.317	1.399	1.398	1.321	1.339	3.6
2-Chlorophenol		1.260	1.297	1.412	1.399	1.359	1.350	4.3
2-Methylphenol		0.915	1.002	1.123	1.140	1.105	1.075	8.0
2,2-oxybis(1-Chloropropane)		1.452	1.496	1.517	1.514	1.405	1.447	4.5
Acetophenone		0.470	0.487	0.522	0.519	0.490	0.495	3.8
3+4-Methylphenols		1.224	1.365	1.560	1.592	1.567	1.491	9.6
n-Nitroso-di-n-propylamine	0.941	1.008	1.008	1.088	1.084	1.041	1.025	4.8
Nitrobenzene-d5		0.322	0.337	0.367	0.368	0.357	0.351	4.8
Hexachloroethane		0.532	0.537	0.558	0.542	0.521	0.533	2.9
Nitrobenzene		0.328	0.334	0.362	0.364	0.349	0.347	4.0
Isophorone		0.575	0.601	0.653	0.673	0.650	0.635	5.4
2-Nitrophenol		0.130	0.143	0.172	0.183	0.184	0.170	14.0
2,4-Dimethylphenol		0.180	0.194	0.219	0.226	0.225	0.213	8.8
bis(2-Chloroethoxy)methane		0.412	0.426	0.444	0.452	0.430	0.429	3.5
2,4-Dichlorophenol		0.242	0.268	0.297	0.308	0.307	0.291	8.9
Naphthalene		1.044	1.054	1.091	1.079	1.037	1.054	2.3
4-Chloroaniline		0.328	0.349	0.389	0.397	0.378	0.371	6.5
Hexachlorobutadiene		0.181	0.181	0.186	0.189	0.181	0.183	1.6
Caprolactam		0.085	0.092	0.111	0.112	0.115	0.107	12.5
4-Chloro-3-methylphenol		0.284	0.310	0.349	0.357	0.356	0.339	8.8
2-Methylnaphthalene		0.700	0.720	0.755	0.755	0.733	0.731	2.9
Hexachlorocyclopentadiene		0.162	0.185	0.197	0.218	0.199	0.195	8.9
2,4,6-Trichlorophenol		0.301	0.335	0.369	0.389	0.386	0.366	9.5
2-Fluorobiphenyl		1.332	1.334	1.359	1.366	1.263	1.306	4.1
2,4,5-Trichlorophenol		0.332	0.371	0.409	0.427	0.429	0.405	9.6
1,1-Biphenyl		1.454	1.489	1.528	1.529	1.434	1.470	3.1
2-Chloronaphthalene		1.063	1.100	1.135	1.139	1.085	1.099	2.6
2-Nitroaniline		0.250	0.270	0.317	0.330	0.324	0.308	11.0
Dimethylphthalate		1.429	1.416	1.495	1.495	1.460	1.454	2.2
Acenaphthylene		1.604	1.680	1.793	1.814	1.726	1.730	4.1

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: SAXT01  
 Lab Code: CHEM Case No.: Q1937 SAS No.: Q1937 SDG No.: Q1937  
 Instrument ID: BNA\_P Calibration Date(s): 04/14/2025 04/14/2025  
 Calibration Time(s): 11:06 17:13

LAB FILE ID:	RRF2.5 = BP024275.D	RRF005 = BP024276.D	RRF010 = BP024277.D	RRF020 = BP024278.D	RRF040 = BP024279.D	RRF050 = BP024280.D	RRF	% RSD
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.264	0.290	0.319	0.325	0.319	0.309	7.6
3-Nitroaniline		0.250	0.289	0.338	0.354	0.342	0.325	12.2
Acenaphthene		1.087	1.087	1.141	1.126	1.081	1.097	2.4
2,4-Dinitrophenol			0.125	0.163	0.186	0.196	0.182	18.2
4-Nitrophenol		0.190	0.231	0.289	0.305	0.306	0.280	17.9
Dibenzofuran		1.801	1.807	1.871	1.833	1.746	1.793	2.8
2,4-Dinitrotoluene		0.334	0.368	0.439	0.448	0.442	0.421	11.8
Diethylphthalate		1.451	1.484	1.528	1.550	1.499	1.499	2.2
4-Chlorophenyl-phenylether		0.667	0.675	0.699	0.680	0.667	0.674	2.0
Fluorene		1.376	1.397	1.454	1.398	1.368	1.388	2.4
4-Nitroaniline		0.263	0.300	0.353	0.369	0.364	0.344	13.0
4,6-Dinitro-2-methylphenol			0.095	0.118	0.131	0.134	0.126	13.5
n-Nitrosodiphenylamine		0.567	0.585	0.616	0.630	0.591	0.597	3.7
2,4,6-Tribromophenol		0.235	0.256	0.278	0.289	0.289	0.277	8.3
4-Bromophenyl-phenylether		0.203	0.212	0.220	0.230	0.220	0.219	4.1
Hexachlorobenzene		0.247	0.247	0.260	0.271	0.261	0.260	3.7
Atrazine		0.176	0.167	0.134	0.121	0.162	0.152	15.6
Pentachlorophenol		0.143	0.155	0.175	0.196	0.197	0.181	13.3
Phenanthrene		1.098	1.086	1.117	1.127	1.070	1.091	2.3
Anthracene		0.995	1.017	1.084	1.111	1.053	1.052	3.9
Carbazole		0.968	0.998	1.060	1.084	1.014	1.027	3.8
Di-n-butylphthalate		1.147	1.230	1.253	1.398	1.310	1.280	6.3
Fluoranthene		1.289	1.274	1.316	1.347	1.280	1.298	2.1
Pyrene		1.192	1.236	1.345	1.305	1.294	1.271	4.4
Terphenyl-d14		0.967	0.998	1.055	1.032	1.005	0.992	4.4
Butylbenzylphthalate		0.434	0.488	0.546	0.596	0.580	0.544	11.2
3,3-Dichlorobenzidine		0.345	0.383	0.431	0.486	0.463	0.436	12.3
Benzo(a)anthracene		1.195	1.214	1.292	1.293	1.234	1.243	3.1
Chrysene		1.180	1.181	1.229	1.213	1.173	1.191	1.9
Bis(2-ethylhexyl)phthalate		0.639	0.735	0.779	0.893	0.853	0.798	11.1
Di-n-octyl phthalate		0.945	1.093	1.222	1.461	1.430	1.297	16.5
Benzo(b)fluoranthene		1.130	1.174	1.217	1.258	1.196	1.199	3.3
Benzo(k)fluoranthene		1.088	1.131	1.217	1.212	1.154	1.158	3.9
Benzo(a)pyrene		0.939	0.971	1.057	1.091	1.053	1.034	5.4
Indeno(1,2,3-cd)pyrene		1.252	1.304	1.407	1.444	1.401	1.388	5.8

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: SAXT01  
 Lab Code: CHEM Case No.: Q1937 SAS No.: Q1937 SDG No.: Q1937  
 Instrument ID: BNA\_P Calibration Date(s): 04/14/2025 04/14/2025  
 Calibration Time(s): 11:06 17:13

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.046	1.089	1.172	1.203	1.169	1.152	5.3
Benzo (g,h,i) perylene		1.073	1.110	1.198	1.218	1.180	1.173	5.0
1,2,4,5-Tetrachlorobenzene		0.522	0.538	0.568	0.576	0.544	0.550	3.3
1,4-Dioxane		0.551	0.517	0.552	0.512	0.481	0.512	6.1
2,3,4,6-Tetrachlorophenol		0.327	0.351	0.380	0.389	0.387	0.373	6.6

All other compounds must meet a minimum RRF of 0.010.