

Surrogate Summary

SDG No.: Q1937

Client: Saxton Falls Sand and Gravel Co. Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1937-02	LARGE-PILE-A	1,2-Dichloroethane-d4	50	54.7	109	70 (63)	130 (155)
		Dibromofluoromethane	50	50.4	101	70 (70)	130 (134)
		Toluene-d8	50	47.6	95	70 (74)	130 (123)
		4-Bromofluorobenzene	50	42.3	85	70 (38)	130 (136)
Q1937-04	LARGE-PILE-B	1,2-Dichloroethane-d4	50	52.5	105	70 (63)	130 (155)
		Dibromofluoromethane	50	48.5	97	70 (70)	130 (134)
		Toluene-d8	50	47.4	95	70 (74)	130 (123)
		4-Bromofluorobenzene	50	39.9	80	70 (38)	130 (136)
Q1937-06	LARGE-PILE-C	1,2-Dichloroethane-d4	50	53.0	106	70 (63)	130 (155)
		Dibromofluoromethane	50	49.7	99	70 (70)	130 (134)
		Toluene-d8	50	48.0	96	70 (74)	130 (123)
		4-Bromofluorobenzene	50	42.0	84	70 (38)	130 (136)
Q1937-08	LARGE-PILE-D	1,2-Dichloroethane-d4	50	52.8	106	70 (63)	130 (155)
		Dibromofluoromethane	50	49.1	98	70 (70)	130 (134)
		Toluene-d8	50	48.0	96	70 (74)	130 (123)
		4-Bromofluorobenzene	50	39.7	79	70 (38)	130 (136)
Q1937-10	LARGE-PILE-E	1,2-Dichloroethane-d4	50	52.2	104	70 (63)	130 (155)
		Dibromofluoromethane	50	49.0	98	70 (70)	130 (134)
		Toluene-d8	50	48.3	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	41.4	83	70 (38)	130 (136)
Q1937-12	LARGE-PILE-F	1,2-Dichloroethane-d4	50	50.9	102	70 (63)	130 (155)
		Dibromofluoromethane	50	49.6	99	70 (70)	130 (134)
		Toluene-d8	50	47.6	95	70 (74)	130 (123)
		4-Bromofluorobenzene	50	36.0	72	70 (38)	130 (136)
VY0502SBL01	VY0502SBL01	1,2-Dichloroethane-d4	50	49.0	98	70 (63)	130 (155)
		Dibromofluoromethane	50	48.7	97	70 (70)	130 (134)
		Toluene-d8	50	47.0	94	70 (74)	130 (123)
		4-Bromofluorobenzene	50	56.8	113	70 (38)	130 (136)
VY0502SBS01	VY0502SBS01	1,2-Dichloroethane-d4	50	50.9	102	70 (63)	130 (155)
		Dibromofluoromethane	50	49.0	98	70 (70)	130 (134)
		Toluene-d8	50	48.6	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.5	101	70 (38)	130 (136)
VY0502SBSD01	VY0502SBSD01	1,2-Dichloroethane-d4	50	52.0	104	70 (63)	130 (155)
		Dibromofluoromethane	50	50.4	101	70 (70)	130 (134)
		Toluene-d8	50	49.7	99	70 (74)	130 (123)
		4-Bromofluorobenzene	50	51.3	103	70 (38)	130 (136)

