

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SAXT01
 Lab Code: CHEM Case No.: Q1937 SAS No.: Q1937 SDG NO.: Q1937
 Lab File ID: VY022119.D Date Analyzed: 05/02/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:56
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	263357	13.347				
UPPER LIMIT	526714	13.847				
LOWER LIMIT	131679	12.847				
EPA SAMPLE NO.						
LARGE-PILE-A	229557	13.35				
LARGE-PILE-B	213475	13.35				
LARGE-PILE-C	222048	13.35				
LARGE-PILE-D	210519	13.35				
LARGE-PILE-E	226299	13.35				
LARGE-PILE-F	170029	13.35				
VY0502SBL01	242655	13.35				
VY0502SBS01	264963	13.35				
VY0502SBSD01	257067	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.