

Initial and Continuing Calibration Verification

Client: Saxton Falls Sand and Gravel Co. Inc.

SDG No.: Q1938

Project: Stan Hope

RunNo.: LB135680

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	05/06/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.490	0.5	98	90-110	05/06/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	05/06/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	05/06/2025

Initial and Continuing Calibration Verification

Client: Saxton Falls Sand and Gravel Co. Inc.

SDG No.: Q1938

Project: Stan Hope

RunNo.: LB135693

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.099	0.099	100	90-110	05/07/2025
Sample ID: CCV1 Cyanide	mg/L	0.25	0.25	100	90-110	05/07/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	05/07/2025
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	05/07/2025
Sample ID: CCV4 Cyanide	mg/L	0.24	0.25	96	90-110	05/07/2025
Sample ID: CCV5 Cyanide	mg/L	0.25	0.25	100	90-110	05/07/2025

Initial and Continuing Calibration Verification

Client: Saxton Falls Sand and Gravel Co. Inc.

SDG No.: Q1938

Project: Stan Hope

RunNo.: LB135693

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------