



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Saxton Falls Sand and Gravel Co. Inc.				SDG No.:	Q1938			
Contract:	SAXT01	Lab Code:	CHEM		Case No.:	Q1938	SAS No.:	Q1938	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB117	Mercury	0.20	+/-0.20	U	0.20	CV	05/06/2025	10:27	LB135678

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc.

SDG No.: Q1938

Contract: SAXT01

Lab Code: CHEM

Case No.: Q1938

SAS No.: Q1938

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB82	Mercury	0.20	+/-0.20	U	0.20	CV	05/06/2025	10:35	LB135678
CCB83	Mercury	0.20	+/-0.20	U	0.20	CV	05/06/2025	11:05	LB135678
CCB84	Mercury	0.20	+/-0.20	U	0.20	CV	05/06/2025	11:32	LB135678
CCB85	Mercury	0.20	+/-0.20	U	0.20	CV	05/06/2025	12:00	LB135678

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc. **SDG No.:** Q1938
Contract: SAXT01 **Lab Code:** CHEM **Case No.:** Q1938 **SAS No.:** Q1938

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	05/07/2025	13:09	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	13:09	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	13:09	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:09	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:09	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	13:09	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	13:09	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	13:09	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	13:09	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	13:09	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	13:09	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	13:09	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	13:09	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	13:09	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	13:09	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	13:09	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	13:09	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	13:09	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	13:09	LB135699

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc. **SDG No.:** Q1938
Contract: SAXT01 **Lab Code:** CHEM **Case No.:** Q1938 **SAS No.:** Q1938

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	05/07/2025	13:43	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	13:43	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	13:43	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:43	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	13:43	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	13:43	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	13:43	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	13:43	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	13:43	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	13:43	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	13:43	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	13:43	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	13:43	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	13:43	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	13:43	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	13:43	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	13:43	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	13:43	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	13:43	LB135699
CCB02	Aluminum	100	+/-100	U	100	P	05/07/2025	14:39	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	14:39	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	14:39	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	14:39	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	14:39	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	14:39	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	14:39	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	14:39	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	14:39	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	14:39	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	14:39	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	14:39	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	14:39	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	14:39	LB135699

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc.		SDG No.: Q1938							
Contract: SEXT01	Lab Code: CHEM	Case No.: Q1938	SAS No.: Q1938						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	14:39	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	14:39	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	14:39	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	14:39	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	14:39	LB135699
CCB03	Aluminum	100	+/-100	U	100	P	05/07/2025	15:27	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	15:27	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	15:27	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	15:27	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	15:27	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	15:27	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	15:27	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	15:27	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	15:27	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	15:27	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	15:27	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	15:27	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	15:27	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	15:27	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	15:27	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	15:27	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	15:27	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	15:27	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	15:27	LB135699
CCB04	Aluminum	100	+/-100	U	100	P	05/07/2025	16:09	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	16:09	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	16:09	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:09	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:09	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	16:09	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	16:09	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	16:09	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	16:09	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	16:09	LB135699

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc. **SDG No.:** Q1938
Contract: SAXT01 **Lab Code:** CHEM **Case No.:** Q1938 **SAS No.:** Q1938

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	16:09	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	16:09	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	16:09	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	16:09	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	16:09	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	16:09	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	16:09	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	16:09	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	16:09	LB135699
CCB05	Aluminum	100	+/-100	U	100	P	05/07/2025	16:52	LB135699
	Antimony	50.0	+/-50.0	U	50.0	P	05/07/2025	16:52	LB135699
	Arsenic	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Barium	100	+/-100	U	100	P	05/07/2025	16:52	LB135699
	Beryllium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:52	LB135699
	Cadmium	6.00	+/-6.00	U	6.00	P	05/07/2025	16:52	LB135699
	Calcium	2000	+/-2000	U	2000	P	05/07/2025	16:52	LB135699
	Chromium	10.0	+/-10.0	U	10.0	P	05/07/2025	16:52	LB135699
	Cobalt	30.0	+/-30.0	U	30.0	P	05/07/2025	16:52	LB135699
	Copper	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Iron	100	+/-100	U	100	P	05/07/2025	16:52	LB135699
	Lead	12.0	+/-12.0	U	12.0	P	05/07/2025	16:52	LB135699
	Magnesium	2000	+/-2000	U	2000	P	05/07/2025	16:52	LB135699
	Manganese	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Nickel	40.0	+/-40.0	U	40.0	P	05/07/2025	16:52	LB135699
	Potassium	2000	+/-2000	U	2000	P	05/07/2025	16:52	LB135699
	Selenium	20.0	+/-20.0	U	20.0	P	05/07/2025	16:52	LB135699
	Silver	10.0	+/-10.0	U	10.0	P	05/07/2025	16:52	LB135699
	Sodium	2000	+/-2000	U	2000	P	05/07/2025	16:52	LB135699
	Thallium	40.0	+/-40.0	U	40.0	P	05/07/2025	16:52	LB135699
	Vanadium	40.0	+/-40.0	U	40.0	P	05/07/2025	16:52	LB135699
	Zinc	40.0	+/-40.0	U	40.0	P	05/07/2025	16:52	LB135699

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc. **SDG No.:** Q1938
Contract: SAXT01 **Lab Code:** CHEM **Case No.:** Q1938 **SAS No.:** Q1938

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	05/09/2025	13:17	LB135724
	Antimony	50.0	+/-50.0	U	50.0	P	05/09/2025	13:17	LB135724
	Arsenic	20.0	+/-20.0	U	20.0	P	05/09/2025	13:17	LB135724
	Barium	100	+/-100	U	100	P	05/09/2025	13:17	LB135724
	Beryllium	6.00	+/-6.00	U	6.00	P	05/09/2025	13:17	LB135724
	Cadmium	6.00	+/-6.00	U	6.00	P	05/09/2025	13:17	LB135724
	Calcium	2000	+/-2000	U	2000	P	05/09/2025	13:17	LB135724
	Chromium	10.0	+/-10.0	U	10.0	P	05/09/2025	13:17	LB135724
	Cobalt	30.0	+/-30.0	U	30.0	P	05/09/2025	13:17	LB135724
	Copper	20.0	+/-20.0	U	20.0	P	05/09/2025	13:17	LB135724
	Iron	100	+/-100	U	100	P	05/09/2025	13:17	LB135724
	Lead	12.0	+/-12.0	U	12.0	P	05/09/2025	13:17	LB135724
	Magnesium	2000	+/-2000	U	2000	P	05/09/2025	13:17	LB135724
	Manganese	20.0	+/-20.0	U	20.0	P	05/09/2025	13:17	LB135724
	Nickel	40.0	+/-40.0	U	40.0	P	05/09/2025	13:17	LB135724
	Potassium	2000	+/-2000	U	2000	P	05/09/2025	13:17	LB135724
	Selenium	20.0	+/-20.0	U	20.0	P	05/09/2025	13:17	LB135724
	Silver	10.0	+/-10.0	U	10.0	P	05/09/2025	13:17	LB135724
	Sodium	2000	+/-2000	U	2000	P	05/09/2025	13:17	LB135724
	Thallium	40.0	+/-40.0	U	40.0	P	05/09/2025	13:17	LB135724
	Vanadium	40.0	+/-40.0	U	40.0	P	05/09/2025	13:17	LB135724
	Zinc	40.0	+/-40.0	U	40.0	P	05/09/2025	13:17	LB135724

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc.		SDG No.: Q1938							
Contract: SEXT01	Lab Code: CHEM	Case No.: Q1938	SAS No.: Q1938						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	05/09/2025	14:15	LB135724
	Antimony	50.0	+/-50.0	U	50.0	P	05/09/2025	14:15	LB135724
	Arsenic	20.0	+/-20.0	U	20.0	P	05/09/2025	14:15	LB135724
	Barium	100	+/-100	U	100	P	05/09/2025	14:15	LB135724
	Beryllium	6.00	+/-6.00	U	6.00	P	05/09/2025	14:15	LB135724
	Cadmium	6.00	+/-6.00	U	6.00	P	05/09/2025	14:15	LB135724
	Calcium	2000	+/-2000	U	2000	P	05/09/2025	14:15	LB135724
	Chromium	10.0	+/-10.0	U	10.0	P	05/09/2025	14:15	LB135724
	Cobalt	30.0	+/-30.0	U	30.0	P	05/09/2025	14:15	LB135724
	Copper	20.0	+/-20.0	U	20.0	P	05/09/2025	14:15	LB135724
	Iron	100	+/-100	U	100	P	05/09/2025	14:15	LB135724
	Lead	12.0	+/-12.0	U	12.0	P	05/09/2025	14:15	LB135724
	Magnesium	2000	+/-2000	U	2000	P	05/09/2025	14:15	LB135724
	Manganese	20.0	+/-20.0	U	20.0	P	05/09/2025	14:15	LB135724
	Nickel	40.0	+/-40.0	U	40.0	P	05/09/2025	14:15	LB135724
	Potassium	2000	+/-2000	U	2000	P	05/09/2025	14:15	LB135724
	Selenium	20.0	+/-20.0	U	20.0	P	05/09/2025	14:15	LB135724
	Silver	10.0	+/-10.0	U	10.0	P	05/09/2025	14:15	LB135724
	Sodium	2000	+/-2000	U	2000	P	05/09/2025	14:15	LB135724
	Thallium	40.0	+/-40.0	U	40.0	P	05/09/2025	14:15	LB135724
	Vanadium	40.0	+/-40.0	U	40.0	P	05/09/2025	14:15	LB135724
	Zinc	40.0	+/-40.0	U	40.0	P	05/09/2025	14:15	LB135724
CCB02	Aluminum	13.6	+/-100	J	100	P	05/09/2025	15:03	LB135724
	Antimony	50.0	+/-50.0	U	50.0	P	05/09/2025	15:03	LB135724
	Arsenic	20.0	+/-20.0	U	20.0	P	05/09/2025	15:03	LB135724
	Barium	100	+/-100	U	100	P	05/09/2025	15:03	LB135724
	Beryllium	6.00	+/-6.00	U	6.00	P	05/09/2025	15:03	LB135724
	Cadmium	6.00	+/-6.00	U	6.00	P	05/09/2025	15:03	LB135724
	Calcium	2000	+/-2000	U	2000	P	05/09/2025	15:03	LB135724
	Chromium	10.0	+/-10.0	U	10.0	P	05/09/2025	15:03	LB135724
	Cobalt	30.0	+/-30.0	U	30.0	P	05/09/2025	15:03	LB135724
	Copper	20.0	+/-20.0	U	20.0	P	05/09/2025	15:03	LB135724
	Iron	100	+/-100	U	100	P	05/09/2025	15:03	LB135724
	Lead	12.0	+/-12.0	U	12.0	P	05/09/2025	15:03	LB135724
	Magnesium	2000	+/-2000	U	2000	P	05/09/2025	15:03	LB135724
	Manganese	20.0	+/-20.0	U	20.0	P	05/09/2025	15:03	LB135724
	Nickel	40.0	+/-40.0	U	40.0	P	05/09/2025	15:03	LB135724
	Potassium	2000	+/-2000	U	2000	P	05/09/2025	15:03	LB135724
	Selenium	20.0	+/-20.0	U	20.0	P	05/09/2025	15:03	LB135724

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc.		SDG No.: Q1938							
Contract: SEXT01	Lab Code: CHEM	Case No.: Q1938	SAS No.: Q1938						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	05/09/2025	15:03	LB135724
	Sodium	2000	+/-2000	U	2000	P	05/09/2025	15:03	LB135724
	Thallium	40.0	+/-40.0	U	40.0	P	05/09/2025	15:03	LB135724
	Vanadium	40.0	+/-40.0	U	40.0	P	05/09/2025	15:03	LB135724
	Zinc	40.0	+/-40.0	U	40.0	P	05/09/2025	15:03	LB135724
CCB03	Aluminum	100	+/-100	U	100	P	05/09/2025	15:58	LB135724
	Antimony	50.0	+/-50.0	U	50.0	P	05/09/2025	15:58	LB135724
	Arsenic	20.0	+/-20.0	U	20.0	P	05/09/2025	15:58	LB135724
	Barium	100	+/-100	U	100	P	05/09/2025	15:58	LB135724
	Beryllium	6.00	+/-6.00	U	6.00	P	05/09/2025	15:58	LB135724
	Cadmium	6.00	+/-6.00	U	6.00	P	05/09/2025	15:58	LB135724
	Calcium	2000	+/-2000	U	2000	P	05/09/2025	15:58	LB135724
	Chromium	10.0	+/-10.0	U	10.0	P	05/09/2025	15:58	LB135724
	Cobalt	30.0	+/-30.0	U	30.0	P	05/09/2025	15:58	LB135724
	Copper	20.0	+/-20.0	U	20.0	P	05/09/2025	15:58	LB135724
	Iron	100	+/-100	U	100	P	05/09/2025	15:58	LB135724
	Lead	12.0	+/-12.0	U	12.0	P	05/09/2025	15:58	LB135724
	Magnesium	2000	+/-2000	U	2000	P	05/09/2025	15:58	LB135724
	Manganese	20.0	+/-20.0	U	20.0	P	05/09/2025	15:58	LB135724
	Nickel	40.0	+/-40.0	U	40.0	P	05/09/2025	15:58	LB135724
	Potassium	2000	+/-2000	U	2000	P	05/09/2025	15:58	LB135724
	Selenium	20.0	+/-20.0	U	20.0	P	05/09/2025	15:58	LB135724
	Silver	10.0	+/-10.0	U	10.0	P	05/09/2025	15:58	LB135724
	Sodium	2000	+/-2000	U	2000	P	05/09/2025	15:58	LB135724
	Thallium	40.0	+/-40.0	U	40.0	P	05/09/2025	15:58	LB135724
	Vanadium	40.0	+/-40.0	U	40.0	P	05/09/2025	15:58	LB135724
	Zinc	40.0	+/-40.0	U	40.0	P	05/09/2025	15:58	LB135724
CCB04	Aluminum	100	+/-100	U	100	P	05/09/2025	16:44	LB135724
	Antimony	50.0	+/-50.0	U	50.0	P	05/09/2025	16:44	LB135724
	Arsenic	20.0	+/-20.0	U	20.0	P	05/09/2025	16:44	LB135724
	Barium	100	+/-100	U	100	P	05/09/2025	16:44	LB135724
	Beryllium	6.00	+/-6.00	U	6.00	P	05/09/2025	16:44	LB135724
	Cadmium	6.00	+/-6.00	U	6.00	P	05/09/2025	16:44	LB135724
	Calcium	2000	+/-2000	U	2000	P	05/09/2025	16:44	LB135724
	Chromium	10.0	+/-10.0	U	10.0	P	05/09/2025	16:44	LB135724
	Cobalt	30.0	+/-30.0	U	30.0	P	05/09/2025	16:44	LB135724
	Copper	20.0	+/-20.0	U	20.0	P	05/09/2025	16:44	LB135724
	Iron	100	+/-100	U	100	P	05/09/2025	16:44	LB135724
	Lead	12.0	+/-12.0	U	12.0	P	05/09/2025	16:44	LB135724

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc.		SDG No.: Q1938							
Contract: SEXT01	Lab Code: CHEM	Case No.: Q1938	SAS No.: Q1938						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	05/09/2025	16:44	LB135724
	Manganese	20.0	+/-20.0	U	20.0	P	05/09/2025	16:44	LB135724
	Nickel	40.0	+/-40.0	U	40.0	P	05/09/2025	16:44	LB135724
	Potassium	2000	+/-2000	U	2000	P	05/09/2025	16:44	LB135724
	Selenium	20.0	+/-20.0	U	20.0	P	05/09/2025	16:44	LB135724
	Silver	10.0	+/-10.0	U	10.0	P	05/09/2025	16:44	LB135724
	Sodium	2000	+/-2000	U	2000	P	05/09/2025	16:44	LB135724
	Thallium	40.0	+/-40.0	U	40.0	P	05/09/2025	16:44	LB135724
	Vanadium	40.0	+/-40.0	U	40.0	P	05/09/2025	16:44	LB135724
	Zinc	40.0	+/-40.0	U	40.0	P	05/09/2025	16:44	LB135724
CCB05	Aluminum	100	+/-100	U	100	P	05/09/2025	17:30	LB135724
	Antimony	50.0	+/-50.0	U	50.0	P	05/09/2025	17:30	LB135724
	Arsenic	20.0	+/-20.0	U	20.0	P	05/09/2025	17:30	LB135724
	Barium	100	+/-100	U	100	P	05/09/2025	17:30	LB135724
	Beryllium	6.00	+/-6.00	U	6.00	P	05/09/2025	17:30	LB135724
	Cadmium	6.00	+/-6.00	U	6.00	P	05/09/2025	17:30	LB135724
	Calcium	2000	+/-2000	U	2000	P	05/09/2025	17:30	LB135724
	Chromium	10.0	+/-10.0	U	10.0	P	05/09/2025	17:30	LB135724
	Cobalt	30.0	+/-30.0	U	30.0	P	05/09/2025	17:30	LB135724
	Copper	20.0	+/-20.0	U	20.0	P	05/09/2025	17:30	LB135724
	Iron	100	+/-100	U	100	P	05/09/2025	17:30	LB135724
	Lead	12.0	+/-12.0	U	12.0	P	05/09/2025	17:30	LB135724
	Magnesium	2000	+/-2000	U	2000	P	05/09/2025	17:30	LB135724
	Manganese	20.0	+/-20.0	U	20.0	P	05/09/2025	17:30	LB135724
	Nickel	40.0	+/-40.0	U	40.0	P	05/09/2025	17:30	LB135724
	Potassium	2000	+/-2000	U	2000	P	05/09/2025	17:30	LB135724
	Selenium	20.0	+/-20.0	U	20.0	P	05/09/2025	17:30	LB135724
	Silver	10.0	+/-10.0	U	10.0	P	05/09/2025	17:30	LB135724
	Sodium	2000	+/-2000	U	2000	P	05/09/2025	17:30	LB135724
	Thallium	40.0	+/-40.0	U	40.0	P	05/09/2025	17:30	LB135724
	Vanadium	40.0	+/-40.0	U	40.0	P	05/09/2025	17:30	LB135724
	Zinc	40.0	+/-40.0	U	40.0	P	05/09/2025	17:30	LB135724
CCB06	Aluminum	100	+/-100	U	100	P	05/09/2025	18:16	LB135724
	Antimony	50.0	+/-50.0	U	50.0	P	05/09/2025	18:16	LB135724
	Arsenic	20.0	+/-20.0	U	20.0	P	05/09/2025	18:16	LB135724
	Barium	100	+/-100	U	100	P	05/09/2025	18:16	LB135724
	Beryllium	6.00	+/-6.00	U	6.00	P	05/09/2025	18:16	LB135724
	Cadmium	6.00	+/-6.00	U	6.00	P	05/09/2025	18:16	LB135724
	Calcium	2000	+/-2000	U	2000	P	05/09/2025	18:16	LB135724

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc.		SDG No.: Q1938							
Contract: SEXT01	Lab Code: CHEM	Case No.: Q1938	SAS No.: Q1938						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	05/09/2025	18:16	LB135724
	Cobalt	30.0	+/-30.0	U	30.0	P	05/09/2025	18:16	LB135724
	Copper	20.0	+/-20.0	U	20.0	P	05/09/2025	18:16	LB135724
	Iron	100	+/-100	U	100	P	05/09/2025	18:16	LB135724
	Lead	12.0	+/-12.0	U	12.0	P	05/09/2025	18:16	LB135724
	Magnesium	2000	+/-2000	U	2000	P	05/09/2025	18:16	LB135724
	Manganese	20.0	+/-20.0	U	20.0	P	05/09/2025	18:16	LB135724
	Nickel	40.0	+/-40.0	U	40.0	P	05/09/2025	18:16	LB135724
	Potassium	2000	+/-2000	U	2000	P	05/09/2025	18:16	LB135724
	Selenium	20.0	+/-20.0	U	20.0	P	05/09/2025	18:16	LB135724
	Silver	10.0	+/-10.0	U	10.0	P	05/09/2025	18:16	LB135724
	Sodium	2000	+/-2000	U	2000	P	05/09/2025	18:16	LB135724
	Thallium	40.0	+/-40.0	U	40.0	P	05/09/2025	18:16	LB135724
	Vanadium	40.0	+/-40.0	U	40.0	P	05/09/2025	18:16	LB135724
	Zinc	40.0	+/-40.0	U	40.0	P	05/09/2025	18:16	LB135724
CCB07	Aluminum	100	+/-100	U	100	P	05/09/2025	18:49	LB135724
	Antimony	50.0	+/-50.0	U	50.0	P	05/09/2025	18:49	LB135724
	Arsenic	20.0	+/-20.0	U	20.0	P	05/09/2025	18:49	LB135724
	Barium	100	+/-100	U	100	P	05/09/2025	18:49	LB135724
	Beryllium	6.00	+/-6.00	U	6.00	P	05/09/2025	18:49	LB135724
	Cadmium	6.00	+/-6.00	U	6.00	P	05/09/2025	18:49	LB135724
	Calcium	2000	+/-2000	U	2000	P	05/09/2025	18:49	LB135724
	Chromium	10.0	+/-10.0	U	10.0	P	05/09/2025	18:49	LB135724
	Cobalt	30.0	+/-30.0	U	30.0	P	05/09/2025	18:49	LB135724
	Copper	20.0	+/-20.0	U	20.0	P	05/09/2025	18:49	LB135724
	Iron	100	+/-100	U	100	P	05/09/2025	18:49	LB135724
	Lead	12.0	+/-12.0	U	12.0	P	05/09/2025	18:49	LB135724
	Magnesium	2000	+/-2000	U	2000	P	05/09/2025	18:49	LB135724
	Manganese	20.0	+/-20.0	U	20.0	P	05/09/2025	18:49	LB135724
	Nickel	40.0	+/-40.0	U	40.0	P	05/09/2025	18:49	LB135724
	Potassium	2000	+/-2000	U	2000	P	05/09/2025	18:49	LB135724
	Selenium	20.0	+/-20.0	U	20.0	P	05/09/2025	18:49	LB135724
	Silver	10.0	+/-10.0	U	10.0	P	05/09/2025	18:49	LB135724
	Sodium	2000	+/-2000	U	2000	P	05/09/2025	18:49	LB135724
	Thallium	40.0	+/-40.0	U	40.0	P	05/09/2025	18:49	LB135724
	Vanadium	40.0	+/-40.0	U	40.0	P	05/09/2025	18:49	LB135724
	Zinc	40.0	+/-40.0	U	40.0	P	05/09/2025	18:49	LB135724

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Saxton Falls Sand and Gravel Co. Inc. **SDG No.:** Q1938
Contract: SAXT01 **Lab Code:** CHEM **Case No.:** Q1938 **SAS No.:** Q1938

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------