

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG NO.: Q1938

Continuing Calib Date: 05/05/2025 Initial Calibration Date(s): 05/02/2025 05/02/2025

Continuing Calib Time: 18:02 Initial Calibration Time(s): 09:45 18:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1248-2	(2)	5.01	5.01	4.91	5.11	0.00
Aroclor-1248-3	(3)	5.23	5.23	5.13	5.33	0.00
Aroclor-1248-4	(4)	5.58	5.58	5.48	5.68	0.00
Aroclor-1248-5	(5)	5.62	5.62	5.52	5.72	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG NO.: Q1938

Continuing Calib Date: 05/05/2025 Initial Calibration Date(s): 05/02/2025 05/02/2025

Continuing Calib Time: 18:02 Initial Calibration Time(s): 09:45 18:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.76	4.76	4.66	4.86	0.00
Aroclor-1248-2	(2)	5.00	5.00	4.90	5.10	0.00
Aroclor-1248-3	(3)	5.04	5.04	4.94	5.14	0.00
Aroclor-1248-4	(4)	5.21	5.21	5.11	5.31	0.00
Aroclor-1248-5	(5)	5.60	5.60	5.50	5.70	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.68	8.68	8.58	8.78	0.00