

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG NO.: Q1938

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL02 Date Analyzed: 05/05/2025

Lab Sample No.: AR1660CCC500 Data File : PO110941.D Time Analyzed: 17:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.774	4.673	4.873	433.660	500.000	-13.3
Aroclor-1016-2	4.793	4.692	4.892	450.100	500.000	-10.0
Aroclor-1016-3	4.850	4.748	4.948	414.720	500.000	-17.1
Aroclor-1016-4	4.970	4.868	5.068	443.310	500.000	-11.3
Aroclor-1016-5	5.227	5.125	5.325	493.700	500.000	-1.3
Aroclor-1260-1	6.267	6.165	6.365	430.420	500.000	-13.9
Aroclor-1260-2	6.456	6.355	6.555	434.950	500.000	-13.0
Aroclor-1260-3	6.823	6.722	6.922	448.030	500.000	-10.4
Aroclor-1260-4	7.084	6.982	7.182	436.580	500.000	-12.7
Aroclor-1260-5	7.326	7.225	7.425	441.680	500.000	-11.7
Decachlorobiphenyl	8.726	8.625	8.825	42.880	50.000	-14.2
Tetrachloro-m-xylene	3.684	3.583	3.783	46.250	50.000	-7.5

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG NO.: Q1938

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL02 Date Analyzed: 05/05/2025

Lab Sample No.: AR1660CCC500 Data File : PO110941.D Time Analyzed: 17:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.759	4.660	4.860	490.500	500.000	-1.9
Aroclor-1016-2	4.778	4.679	4.879	495.300	500.000	-0.9
Aroclor-1016-3	4.953	4.854	5.054	497.660	500.000	-0.5
Aroclor-1016-4	4.995	4.896	5.096	468.240	500.000	-6.4
Aroclor-1016-5	5.208	5.109	5.309	538.160	500.000	7.6
Aroclor-1260-1	6.238	6.139	6.339	487.300	500.000	-2.5
Aroclor-1260-2	6.426	6.327	6.527	485.210	500.000	-3.0
Aroclor-1260-3	6.579	6.480	6.680	488.250	500.000	-2.4
Aroclor-1260-4	7.049	6.950	7.150	483.810	500.000	-3.2
Aroclor-1260-5	7.290	7.192	7.392	489.070	500.000	-2.2
Decachlorobiphenyl	8.675	8.576	8.776	49.190	50.000	-1.6
Tetrachloro-m-xylene	3.680	3.580	3.780	50.860	50.000	1.7