

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG NO.: Q1938

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP071781.D Time Analyzed: 16:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.665	5.571	5.771	516.250	500.000	3.3
Aroclor-1016-2	5.686	5.593	5.793	519.410	500.000	3.9
Aroclor-1016-3	5.748	5.654	5.854	517.560	500.000	3.5
Aroclor-1016-4	5.845	5.752	5.952	525.690	500.000	5.1
Aroclor-1016-5	6.138	6.045	6.245	514.100	500.000	2.8
Aroclor-1260-1	7.257	7.165	7.365	494.870	500.000	-1.0
Aroclor-1260-2	7.511	7.418	7.618	498.270	500.000	-0.3
Aroclor-1260-3	7.869	7.777	7.977	512.800	500.000	2.6
Aroclor-1260-4	8.093	8.002	8.202	493.240	500.000	-1.4
Aroclor-1260-5	8.412	8.322	8.522	518.950	500.000	3.8
Decachlorobiphenyl	10.224	10.137	10.337	53.190	50.000	6.4
Tetrachloro-m-xylene	4.511	4.417	4.617	51.380	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Contract: SAXT01

Lab Code: CHEM Case No.: Q1938 SAS No.: Q1938 SDG NO.: Q1938

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP071781.D Time Analyzed: 16:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.890	4.798	4.998	512.360	500.000	2.5
Aroclor-1016-2	4.909	4.816	5.016	520.640	500.000	4.1
Aroclor-1016-3	5.086	4.993	5.193	540.620	500.000	8.1
Aroclor-1016-4	5.127	5.036	5.236	545.370	500.000	9.1
Aroclor-1016-5	5.342	5.250	5.450	544.720	500.000	8.9
Aroclor-1260-1	6.376	6.285	6.485	540.930	500.000	8.2
Aroclor-1260-2	6.564	6.473	6.673	546.000	500.000	9.2
Aroclor-1260-3	6.716	6.627	6.827	549.500	500.000	9.9
Aroclor-1260-4	7.188	7.098	7.298	552.720	500.000	10.5
Aroclor-1260-5	7.429	7.339	7.539	547.720	500.000	9.5
Decachlorobiphenyl	8.834	8.749	8.949	54.930	50.000	9.9
Tetrachloro-m-xylene	3.807	3.712	3.912	53.670	50.000	7.3