



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1938

Client: Saxton Falls Sand and Gravel Co. Inc.

Analytical Method: 8082A Datafile : PO110932.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB167853BS	AR1016	166.6	148	ug/kg	89				40 (71)	140 (120)		
	AR1260	166.6	146	ug/kg	88				40 (65)	140 (130)		